

Quasimeme Laboratory Performance Studies



Round 2017 - 1

3 April 2017 to 1 July 2017

Exercise Protocols

Version 1: 29th March, 2017

Table of Contents

Introduction	Round 2017 - 1	4
AQ-1	Nutrients in Seawater	6
AQ-2	Nutrients in Estuarine and Low Salinity Open Seawater	8
AQ-14	DOC in seawater	10
AQ-3	Metals in Seawater	12
AQ-4	Mercury in Seawater	14
AQ-5	Halogenated Organics in Seawater	16
AQ-6	Volatile Organics in Seawater	18
AQ-7	Pentachlorophenol in Seawater	20
AQ-8	Triazines and Organophosphorus Compounds in the Seawater	22
AQ-11	Chlorophyll-a in Seawater	24
AQ-12	Organotins in Seawater	26
AQ-13	Polycyclic Aromatic Hydrocarbons in Seawater	28
MS-1	Trace metals in Sediment	30
MS-2	Chlorinated Organics in Sediment	32
MS-3	Polycyclic Aromatic Hydrocarbons in Sediment	34
MS-6	Organotins in Sediment	36
MS-7	Brominated flame retardants in Sediment	38
MS-8	PFASs in Sediment	40
BT-1	Trace metals in Biota	42
BT-2	Chlorinated Organics in Biota	44
BT-4	Polycyclic Aromatic Hydrocarbons in Biota	47
BT-8	Organotins in Biota	49
BT-9	Brominated Flame Retardants in Biota	51
BT-10	Perfluorinated Alkyl Substances (PFASs) in Biota	53
BT-7	ASP Shellfish Toxins	55
BT-11	Lipophilic Shellfish Toxins	57

BT-12 PSP Shellfish Toxins 60

Reporting of Results and Analytical Methods 63

ANNEX 1 Notification of damaged test materials..... 65

ANNEX 2 Instructions for login into Participant Site 66

ANNEX 3 Total Lipid Extraction According to Smedes 67

Introduction Round 2017 - 1

Thank you for participating in the 2017 QUASIMEME Laboratory Performance studies.

The test materials for the exercises in Round 2017-1, that you have ordered will be sent to you by courier in the week of 3 April 2017. Please check that the contents of your package are correct and that all test materials are intact. If any test materials have been damaged in transit or if the wrong samples have been send, use the form in Annex 1 of this document to request replacement materials within two weeks after receipt of the test materials.

Additional test materials may also be purchased from QUASIMEME.

This protocol covers the following studies :

Round	Analysis Group Code	Matrix	Analytes
2017 - 1	AQ-1	Seawater	Nutrients
2017 - 1	AQ-2	Estuarine and Low Salinity Open Water	Nutrients
2017 - 1	AQ-3	Seawater	Metals
2017 - 1	AQ-4	Seawater	Mercury
2017 - 1	AQ-5	Seawater	Halogenated Organics
2017 - 1	AQ-6	Seawater	Volatile Organics
2017 - 1	AQ-7	Seawater	Pentachlorophenol
2017 - 1	AQ-8	Seawater	Triazines and Organophosphorus Compounds
2017 - 1	AQ-11	Seawater	Chlorophyll and Pheopigments
2017 - 1	AQ-12	Seawater	Organotins
2017 - 1	AQ-13	Seawater	Polycyclic Aromatic Hydrocarbons
2017 - 1	AQ-14	Seawater	DOC
2017 - 1	MS-1	Sediment	Trace Metals
2017 - 1	MS-2	Sediment	Chlorinated Organics
2017 - 1	MS-3	Sediment	Polycyclic Aromatic Hydrocarbons
2017 - 1	MS-6	Sediment	Organotins
2017 - 1	MS-7	Sediment	Brominated Flame Retardants
2017 - 1	MS-8	Sediment	PFASs in sediment
2017 - 1	BT-1	Biota	Trace Metals
2017 - 1	BT-2	Biota	Chlorinated Organics
2017 - 1	BT-4	Biota	Polycyclic Aromatic Hydrocarbons
2017 - 1	BT-8	Biota	Organotins
2017 - 1	BT-9	Biota	Brominated Flame Retardants
2017 - 1	BT-10	Biota	Perfluorinated Alkyl Substances (PFASs)
2017 - 1	BT-7	Shellfish and Solution	ASP Shellfish Toxins
2017 - 1	BT-11	Shellfish and Solution	DSP Shellfish Toxins
2017 - 1	BT-12	Shellfish	PSP Shellfish Toxins

All data for these studies must be uploaded to your Quasimeme SharePoint Site, using the Data Submission Forms, no later than 1st July 2017.

IMPORTANT: Please note that the Data Submission Forms are changed, report **ONLY** the parameters as they are asked in this protocol

All other information should be sent to: QUASIMEME Project Office

<p>Wageningen University & Research WEPAL-QUASIMEME Project Office P.O. Box 8005 6700 EC Wageningen The Netherlands</p> <p>Bornesesteeg 10 6721 NG Bennekom The Netherlands</p>	<p>Website: http://www.Quasimeme.org Tel.: +31 (0) 317 48 65 46 Fax: +31 (0) 317 48 56 66 E-mail: Quasimeme@wur.nl</p>
---	---

ROUND	2017 - 1
AQ-1	Nutrients in Seawater
Test materials	QNU305SW, QNU306SW and QNU307SW

Objective

This study covers the determination of nutrients in the seawater test materials.

Test Materials and storage

The test materials were prepared at the laboratory of RBINS (Royal Belgium Institute of Natural Sciences), Ostend, Belgium, using seawater collected from the North Sea between Belgium and the UK.

The seawater was filtered using a 0.45µm / 0.2µm double membrane filter. The pH of the filtered seawater was adjusted to ~ pH 7.2 with 0.1M hydrochloric acid and spiked to appropriate concentrations. The spiked seawater is then thoroughly mixed before being dispensed into the glass and plastic bottles. The filled bottles are then autoclaved at 110°C, 1.5 bar for 30 minutes. The autoclaving process removes micro-organisms, which affect the stability of the nutrient test materials. It has been demonstrated that autoclaving the test materials generates an increase in pH. We have found that after autoclaving, the pH of the nutrient test materials is within the range pH 7.5 to 8.5.

The three test materials differ from each other in respect of their nutrient concentrations.

There are two bottles for each test material - one glass and one plastic. The glass bottle should only be used for the determination of TOxN, nitrite, ammonia and total-N. The plastic bottle should only be used for the determination of silicate, phosphate and total-P. Each bottle contains approximately 250 ml of the test material .

There is a separate bottle for the determination of salinity labelled Salinity ONLY (QNU305SW). Salinity should NOT be measured and/or reported for both other samples.

Each batch of material was prepared in bulk. Homogeneity testing is performed on each batch of test materials produced. The nutrient test materials are stable for the period of the test, and have also been shown to be stable for a period of some months, even after opening, if used under the correct conditions.

Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Treat all test materials in the same manner as your routine samples.

Samplecode	Description
QNU305SW	Seawater (Salinity > 30 psu)
QNU306SW	Seawater (Salinity > 30 psu) spiked
QNU307SW	Seawater (Salinity > 30 psu) spiked

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following nutrients should be determined:

Determinand	Unit	Concentration range		Error		AA-EQS
		Seawater	Seawater (spiked)	Const	Prop	
Ammonia	µmol/L	0.2–5	0.2–5	0.1	6.0%	
Nitrite	µmol/L	0.01–2	0.01–2	0.01	6.0%	
Phosphate	µmol/L	0.05–5	0.05–5	0.05	6.0%	
Silicate	µmol/L	0.5–10	0.5–10	0.1	6.0%	
Total-N	µmol/L	5–25	5–25	0.5	6.0%	
Total-P	µmol/L	0.1–5	0.1–5	0.05	6.0%	
TOxN	µmol/L	0.05–15	0.05–15	0.05	6.0%	
Salinity	psu			0.01	0.1%	

Please report Salinity only for QNU305SW

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material “as received”. The concentration of nutrients should be determined against your own calibration solutions.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 -1
AQ-2 Nutrients in Estuarine and Low Salinity Open Seawater	
Test materials	QNU308EW, QNU309EW, QNU310EW and QNU311EW

Objective

This study covers the determination of nutrients in estuarine water test materials and low salinity open water test materials.

Test Materials and storage

The test materials were prepared at the laboratory of RBINS (Royal Belgium Institute of Natural Sciences), Ostend, Belgium, using seawater collected from the North Sea between Belgium and the UK, and from the Baltic Sea (low salinity water samples)

The seawater was filtered using a 0.45µm / 0.2µm double-membrane filter and diluted with ultrapure demineralised water. The pH of the filtered seawater was adjusted to ~pH7.2 with 0.1M hydrochloric acid and spiked to appropriate concentrations. The spiked seawater is then thoroughly mixed before being dispensed into the glass and plastic bottles. The filled bottles are then autoclaved at 110°C, 1.5 bar for 30 minutes. The autoclaving process removes micro-organisms, which affect the stability of the nutrient test materials. It has been demonstrated that autoclaving the test materials generates an increase in pH. We have found that after autoclaving, the pH of the nutrient test materials is within the range pH 7.5 to 8.5. The four test materials differ from each other in respect of their nutrient concentrations and the salinity of the water. The salinity of the water will be approximately 8-20 psu. One of the samples is the unspiked sample and the other samples are spiked with nutrients.

There are two bottles for each test material - one glass and one plastic. The glass bottle should only be used for the determination of TOxN, nitrite, ammonia and total-N. The plastic bottle should only be used for the determination of silicate, phosphate and total-P.

There is a separate bottle for the determination of salinity labelled Salinity ONLY (QNU308SW). Salinity should NOT be measured and/or reported for the other samples.

Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Treat all test materials in the same manner as your routine samples.

Samplecode	Description
QNU308EW	Estuarine water (Salinity 8 - 15 psu) spiked
QNU309EW	Estuarine water (Salinity 8 - 15 psu) spiked
QNU310EW	Low salinity seawater (Salinity 10 - 20 psu) spiked
QNU311EW	Unspiked Low salinity seawater (Salinity 10 - 20 psu)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following nutrients should be determined:

Determinand	Unit	Concentration range		Error		AA-EQS
		Estuarine water (spiked)	Low salinity open water (spiked)	Const	Prop	
Ammonia	µmol/L	2–50	0.2–5	0.1	6.0%	
Nitrite	µmol/L	0.5–25	0.01–2	0.01	6.0%	
Phosphate	µmol/L	1–15	0.02–5	0.05	6.0%	
Salinity	psu			0.01	0.1%	
Silicate	µmol/L	5–100	0.5–20	0.1	6.0%	
Total-N	µmol/L	10–200	2–20	0.5	6.0%	
Total-P	µmol/L	1–20	0.02–2	0.05	6.0%	
TOxN	µmol/L	10–100	0.05–15	0.05	6.0%	

Please report Salinity only for QNU308EW

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Only one result for the different nutrients per test material is required. The results should be expressed on the test material “as received”. The concentration of the nutrients should be determined against your own calibration solutions.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-14	DOC in seawater
Test materials	QDC049SW, QDC050SW, QDC051EW and QDC052EW

Objective

This study covers the determination of DOC in open seawater and estuarine water test materials.

Test Materials and storage

The test materials were prepared at the laboratory of RBINS (Royal Belgium Institute of Natural Sciences), Ostend, Belgium, using seawater collected from the North Sea between Belgium and the UK.

The seawater was filtered using a 0.45µm / 0.2µm double-membrane filter and diluted with ultrapure demineralised water for the estuarial samples. The pH of the filtered seawater was adjusted to ~pH7.2 with 0.1M hydrochloric acid and spiked to appropriate concentrations. The spiked seawater is then thoroughly mixed before being dispensed into the glass bottles. The filled bottles are then autoclaved at 110°C, 1.5 bar for 30 minutes. The autoclaving process removes micro-organisms, which affect the stability of the DOC test materials. It has been demonstrated that autoclaving the test materials generates an increase in pH. We have found that after autoclaving, the pH of the DOC test materials is within the range pH 7.5 to 8.5. The four test materials differ from each other in respect of their DOC concentrations and the salinity of the water. The salinity of the estuarine seawater will be approximately 7 - 10 psu. One of the samples is the unspiked sample and the other samples are spiked with DOC.

Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QDC049SW	Seawater (Salinity > 30 psu)
QDC050SW	Seawater (Salinity > 30 psu) spiked
QDC051EW	Estuarine water (Salinity 7 - 10 psu) spiked
QDC052EW	Estuarine water (Salinity 7 - 10 psu) spiked

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following nutrients should be determined:

Determinand	Unit	Concentration range			Error		AA-EQS
		Seawater	Seawater (spiked)	Estuarine water (spiked)	Const	Prop	
DOC	mg C/L	0.1–20	0.1–20	0.1–20	0.1	6.0%	

Analysis

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Only one result for DOC test material is required. The result should be expressed on the test material "as received". The concentration of DOC should be determined against your own calibration solutions.

Reporting

The result for DOC should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-3	Metals in Seawater
Test materials	QTM247SW, QTM248SW, QTM249SW and QTM250SW

Objective

This study covers the determination of metals in seawater and low salinity seawater test materials.

Test Materials and storage

The test materials were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK.

The test materials were prepared in bulk in a 50 litre vessel. The seawater was filtered using a 0.45 µm /0.2µm double-membrane filter. Low salinity seawater test material was prepared by diluting the seawater with ultra-pure demineralised water. All test materials are preserved with 2 ml trace metal grade nitric acid per litre of seawater. Spiked test materials were prepared by adding aqueous solutions of known trace metal concentration. Approximately 1 litre of each test material is provided. Homogeneity of the test materials is assumed, as they were prepared in bulk and thoroughly mixed, before being dispensed into 1 litre polyethylene bottles. The test materials are stable for the purposes of the exercise.

Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Samplecode	Description
QTM247SW	Unspiked Seawater (Salinity > 30 psu)
QTM248SW	Seawater (Salinity > 30 psu) spiked with metals
QTM249SW	Low salinity Seawater (Salinity 8 - 20 psu) spiked with metals
QTM250SW	Low salinity Seawater (Salinity 8 -20 psu) sample spiked with concentrations between 5 and 50 times higher than the indicative range (500 ml)

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Analysis

Treat all test materials in the same manner as your routine samples.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Only one result per determinand per test material is required.

Determinands and concentration ranges

The following metals should be determined:

Determinand	Unit	Concentration range		Error		AA-EQS
		Seawater (spiked)	Low salinity Seawater (spiked)	Const	Prop	
Arsenic	µg/L	0.05–5	0.2–10	0.5	12.5%	
Boron	µg/L	1000–5000	200–5000	0.4	12.5%	
Cadmium	µg/L	0.001–0.5	0.05–1	0.005	12.5%	0.2
Chromium	µg/L	0.01–5	0.5–10	0.1	12.5%	
Cobalt	µg/L	0.001–0.5	0.01–10	0.2	12.5%	
Copper	µg/L	0.05–10	0.2–10	0.2	12.5%	
Iron	µg/L	0.05–10	0.2–20	0.4	12.5%	
Lead	µg/L	0.0002–15	0.1–5	0.01	12.5%	7.2
Manganese	µg/L	0.02–5	0.1–5	0.4	12.5%	
Nickel	µg/L	0.2–5	0.1–5	0.2	12.5%	20
Silver	µg/L	0.02–2	0.1–5	0.2	12.5%	
Tin	µg/L	0.02–1	0.1–5	0.2	12.5%	
Vanadium	µg/L	0.1–5	0.2–5	0.2	12.5%	
Zinc	µg/L	0.5–20	0.2–20	0.4	12.5%	

QTM250SW contains concentrations of the determinands which are 5 to 50 times higher compared to the indicative range given in this protocol.

Boron is naturally occurring at higher concentrations.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-4	Mercury in Seawater
Test materials	QTM251SW, QTM252SW, QTM253SW and QTM254SW

Objective

This study covers the determination of mercury in the seawater test materials. The test materials should be analysed and one result for mercury in each test material should be reported using the Data Submission Forms provided on the Participant Site.

Test Materials and storage

The test materials were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK.

The test materials were prepared in bulk in a 50 litre vessel. The seawater was filtered using a 0.45 µm / 0.2 µm double-membrane filter. All test materials are preserved with 2 ml trace metal grade nitric acid per litre of seawater. Test materials were spiked with aqueous solutions of known mercury concentration.

Approximately 1 litre of each test material is provided.

Homogeneity of the test materials is assumed, as they were prepared in bulk and thoroughly mixed, before being dispensed into 1 litre glass bottles. The test materials are stable for the purposes of the exercise. Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Samplecode	Description
QTM251SW	Seawater (Salinity > 30 psu) spiked with mercury
QTM252SW	Seawater (Salinity > 30 psu) spiked with mercury
QTM253SW	Low salinity Seawater (Salinity 10 - 20 psu) spiked with mercury
QTM254SW	Low salinity Seawater (Salinity 10 - 20 psu) spiked sample with concentrations between 5 and 50 times higher compared to the concentrations given in this protocol

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

Mercury should be determined in each test material.

Determinand	Unit	Concentration range	Error		AA-EQS
		Seawater (spiked)	Const	Prop	
Mercury	ng/L	0.1—100	0.2	12.5%	50

QTM254SW contains concentrations of the determinands which are 5 to 50 times higher compared to the indicative range given in this protocol.

Analysis

Treat all test materials in the same manner as your routine samples.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. Only one result per test material is required.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-5 Halogenated Organics in Seawater	
Test materials	QOC088SW, QOC089SW, QOC090SW and QOC091SW

Objective

This study covers the determination of Halogenated organics in seawater test materials.

Test Materials and storage

The test materials were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK. The test materials were prepared in bulk in a 100 litre vessel. The seawater was filtered using a 0.45 µm / 0.2 µm double-membrane filter. Low salinity seawater test material are prepared by dilution with ultra-pure demineralised water, to a salinity of approximately 12 - 18 psu. Test materials need to be spiked with organochlorine composite solutions in methanol by the participants themselves (see Analysis section). Homogeneity of the test materials is assumed, as they were prepared in bulk and thoroughly mixed, before being dispensed into 1 litre glass bottles. The test materials are stable for the purposes of the exercise. Approximately 1 litre of each test material is provided. Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Treat all test materials in the same manner as your routine samples.

Sample code	Description
QOC088SW	Seawater with Spiking solution QOC088SS (low level: 0-10 ng/L)
QOC089SW	Seawater with Spiking solution QOC089SS (mid level: 1 -25 ng/L)
QOC090SW	Low salinity Seawater with Spiking solution QOC090SS (high level: 2,5 - 150 ng/L)
QOC091SW	Low salinity Seawater with Spiking solution QOC091SS (very high level: 50-2000 ng/L)

N.B. Please use the correct spiking solution with the correct bottle of seawater. Use the spiking solution labelled 'QOC088SS spiking solution' ONLY with the seawater bottle labelled QOC088SW.

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following organochlorine compounds should be determined:

Determinand	Unit	Concentration Range		Error		AA-EQS
		Low Salinity Seawater (spiked)	Seawater (spiked)	Const	Prop	
□-HCH	ng/L	2–50	0.2–20	0.2	12.5%	2
□-HCH	ng/L	1–50	0.2–20	0.2	12.5%	2
□-HCH	ng/L	2–50	0.5–20	0.2	12.5%	2
□-HCH	ng/L	1–50	0.2–20	0.2	12.5%	2
HCB	ng/L	0.5–20	0.1–10	0.2	12.5%	10
HCBD	ng/L	2–50	0.2–20	0.2	12.5%	100
Aldrin	ng/L	2–200	1–20	0.5	12.5%	5
Dieldrin	ng/L	2–100	1–20	0.5	12.5%	5
Endrin	ng/L	2–200	1–20	0.5	12.5%	5
Isodrin	ng/L	2–200	1–20	0.5	12.5%	5

pp'-DDD	ng/L	1–50	0.1–10	0.5	12.5%	25
pp'-DDE	ng/L	1–50	0.2–10	0.5	12.5%	25
op'-DDT	ng/L	1–50	0.2–20	0.5	12.5%	25
pp'-DDT	ng/L	1–50	0.2–20	0.5	12.5%	10
Endosulphan-I	ng/L	1–20	0.2–10	0.2	12.5%	0.5
Endosulphan-II	ng/L	0.5–20	0.1–10	0.2	12.5%	0.5
Pentachlorobenzene	ng/L	2–100	0.2–5	0.5	12.5%	0.7
1,2,3-TCB	ng/L	2–50	1–20	0.5	12.5%	400
1,2,4-TCB	ng/L	5–100	1–20	0.5	12.5%	400
1,3,5-TCB	ng/L	2–50	0.5–20	0.5	12.5%	400
Trifluralin	ng/L	2–50	0.5–20	0.5	12.5%	30
PCB28	ng/L	2–50	0.5–20	0.2	12.5%	
PCB31	ng/L	2–50	0.5–20	0.2	12.5%	
PCB52	ng/L	2–50	0.5–20	0.2	12.5%	
PCB101	ng/L	2–50	0.5–20	0.2	12.5%	
PCB105	ng/L	2–50	0.5–20	0.2	12.5%	
PCB118	ng/L	2–50	0.5–20	0.2	12.5%	
PCB138	ng/L	2–50	0.5–20	0.2	12.5%	
PCB138 + PCB163	ng/L	2–50	0.5–20	0.2	12.5%	
PCB153	ng/L	2–50	0.5–20	0.2	12.5%	
PCB156	ng/L	2–50	0.5–20	0.2	12.5%	
PCB180	ng/L	2–50	0.5–20	0.2	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

Preparation and Analysis

A 2000-times dilution of the spiking solutions is required, using the seawater test materials to produce the spiked seawater (E.g. 0.8 litre seawater + 400 µl spiking solution). The dilution procedure is given below:

The spiking solution should be stabilised at room temperature

Weigh an empty flask, which is normally used within your laboratory for this kind of analysis. Weigh the required amount of the seawater provided, in the flask. A mass of 820 gram of water is equal to 800 ml of seawater (density=1.025 kg/L).

Add the spiking solution to the flask with seawater, corresponding to a 2000-times dilution (e.g. 0.8 litre of seawater + 400 µl of spiking solution) and mix thoroughly. You may use different volumes but the dilution factor should always be kept to a factor of 2000.

Analyse the test materials immediately after preparation. Realize that some determinands can adsorb to the wall of the flask due to limited solubility in seawater of those determinands.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-6 Volatile Organics in Seawater	
Test materials	QVC061SW and QVC062SW

Objective

This study covers the determination of volatile organochlorine compounds in seawater test materials.

Test Materials and storage

The test materials were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK, and was stored in the cold store at 7 °C in 25 litre carboys. The test materials were prepared in bulk in a 100 litre vessel. The seawater was filtered using a 0.45 µm / 0.2 µm double-membrane filter. Test materials were spiked with the volatile organochlorine composite solution in methanol with known concentration. Flasks were completely filled with test material. Homogeneity of the test materials is assumed, as they were prepared in bulk and thoroughly mixed, before being dispensed into 1 litre glass bottles. The test materials are stable for the purposes of the exercise. Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Treat all test materials in the same manner as your routine samples.

Samplecode	Description
QVC061SW	Seawater (Salinity > 30 psu) spiked
QVC062SW	Seawater (Salinity > 30 psu) spiked

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Seawater (spiked)	Const	Prop	
1,1,1-Trichloroethane	µg/L	0.2–10	0.1	12.5%	
1,1,2-Trichloroethane	µg/L	1–20	0.1	12.5%	
1,2-Dichloroethane	µg/L	1–10	0.1	12.5%	10
Benzene	µg/L	0.2–50	0.1	12.5%	8
Carbontetrachloride	µg/L	0.2–10	0.1	12.5%	12
Chloroform	µg/L	0.5–20	0.1	12.5%	2.5
Dichloromethane	µg/L	0.2–20	0.1	12.5%	20
Tetrachloroethene	µg/L	0.2–10	0.1	12.5%	10
Trichloroethene	µg/L	0.2–10	0.1	12.5%	10
Styrene	µg/L	0.1–50	0.1	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material "as received". The concentration of the volatiles should be determined against your own calibration solutions.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-7	Pentachlorophenol in Seawater
Test materials	QPP063SW, QPP064SW and QPP065SW

Objective

This study covers the determination of Pentachlorophenol in seawater test materials.

Test Materials and storage

The test materials were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK, and was stored in the cold store at 7 °C in 25 litre carboys. The test materials were prepared in bulk in a 100 litre vessel. The seawater was filtered using a 0.45 µm / 0.2 µm double-membrane filter. Test materials need to be spiked with pentachlorophenol solutions in methanol by the participants themselves (see Analysis section). Homogeneity of the test materials is assumed, as they were prepared in bulk and thoroughly mixed, before being dispensed into 1 litre glass bottles. The test materials are stable for the purposes of the exercise. Approximately 1 litre of each test material is provided. Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Treat all test materials in the same manner as your routine samples.

Samplecode	Description
QPP063SW	Seawater with Spiking solution QPP063SS
QPP064SW	Seawater with Spiking solution QPP064SS
QPP065SW	Seawater with Spiking solution QPP065SS

N.B. Please use the correct spiking solution with the correct bottle of seawater. Use the spiking solution labelled 'QPP063SS spiking solution' ONLY with the seawater bottle labelled QPP063SW.

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

Pentachlorophenol should be determined in each test material.

Determinand	Unit	Concentration range	Error		AA-EQS
			Const	Prop	
Pentachlorophenol	ng/L	20-2000	10	12.5%	400

Analysis

Preparation and Analysis

A 2000-times dilution of the spiking solutions is required, using the seawater test materials to produce the spiked seawater (E.g. 0.8 litre seawater + 400 µl spiking solution). The dilution procedure is given below:

The spiking solution should be stabilised at room temperature

Weigh an empty flask, which is normally used within your laboratory for this kind of analysis. Weigh the required amount of the seawater provided, in the flask. A mass of 820 gram of water is equal to 800 ml of seawater (density=1.025 kg/L).

Add the spiking solution to the flask with seawater, corresponding to a 2000-times dilution (e.g. 0.8 litre of seawater + 400 µl of spiking solution) and mix thoroughly. You may use different volumes but the dilution factor should always be kept to a factor of 2000.

Analyse the test materials immediately after preparation. Realize that some determinands can adsorb to the wall of the flask due to limited solubility in seawater of those determinands.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Reporting

One result for each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-8 Triazines and Organophosphorus Compounds in the Seawater	
Test materials	QTP094SW, QTP095SW and QTP096SW

Objective

This study covers the determination of triazines and organophosphorus compounds in the seawater.

Test Materials and storage

The test materials were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK, and was stored in the cold store at 7 °C in 25 litre carboys. The test materials were prepared in bulk in a 100 litre vessel. The seawater was filtered using a 0.45 µm / 0.2 µm double-membrane filter.

Methanol solutions containing known concentrations of organophosphorus compounds and triazines were prepared in bulk and ampouled to make the spiking solutions.

For each test material, approximately 1 litre of filtered seawater and an ampoule of spiking solution is provided.

Homogeneity of the test materials is assumed, as they were prepared from the same bulk seawater, and the spiking solutions were also prepared in bulk. The test materials are stable for the purposes of the exercise. Test materials (seawater and spiking solutions) should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately. Treat all test materials in the same manner as your routine samples.

Samplecode	Description
QTP094SW	Seawater with Spiking solution QTP094SS (low level: 2-20 ng/L)
QTP095SW	Seawater with Spiking solution QTP095SS (mid level: 20-200 ng/L)
QTP096SW	Low salinity Seawater with Spiking solution QTP096SS (high level: 50-500 ng/L)

*N.B. Please use the correct spiking solution with the correct bottle of seawater. Use the spiking solution labelled QTP094SS spiking solution **ONLY** with the bottle of seawater labelled QTP094SW.*

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range		Error		AA-EQS
		Seawater with SS	Low salinity Seawater with SS	Const	Prop	
Alachlor	ng/L	2–200	20–500	1	12.5%	300
Atrazine	ng/L	5–200	20–500	1	12.5%	600
Azinphos-ethyl	ng/L	5–200	20–500	1	12.5%	
Azinphos-methyl	ng/L	5–200	20–500	1	12.5%	
Chlorfenvinphos	ng/L	5–200	20–500	1	12.5%	100
Chlorpyriphos	ng/L	2–200	20–500	1	12.5%	30
Coumaphos	ng/L	2–100	20–500	1	12.5%	
Demeton	ng/L	5–200	50–500	1	12.5%	
Diazinon	ng/L	5–200	20–500	1	12.5%	
Dichlorvos	ng/L	2–200	20–500	1	12.5%	

Dimethoate	ng/L	5–100	20–500	1	12.5%	
Diuron	ng/L	5–200	50–500	1	12.5%	200
Fenclorophos	ng/L	2–200	20–500	1	12.5%	
Fenitrothion	ng/L	2–200	20–500	1	12.5%	
Fenthion	ng/L	5–200	20–500	1	12.5%	
Irgarol-1051	ng/L	2–200	50–500	1	12.5%	
Isoproturon	ng/L	2–200	20–500	1	12.5%	300
Malathion	ng/L	5–200	20–500	1	12.5%	
Omethoate	ng/L	5–200	50–500	1	12.5%	
Parathion-ethyl	ng/L	5–200	20–500	1	12.5%	
Parathion-methyl	ng/L	5–200	20–500	1	12.5%	
Simazine	ng/L	5–200	20–500	1	12.5%	1000
Triazophos	ng/L	10–500	50–500	1	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

Preparation and Analysis

A 2000-times dilution of the spiking solutions is required, using the seawater test materials to produce the spiked seawater (E.g. 0.8 litre seawater + 400 µl spiking solution). The dilution procedure is given below:

The spiking solution should be stabilised at room temperature

Weigh an empty flask, which is normally used within your laboratory for this kind of analysis. Weigh the required amount of the seawater provided, in the flask. A mass of 820 gram of water is equal to 800 ml of seawater (density=1.025 kg/L).

Add the spiking solution to the flask with seawater, corresponding to a 2000-times dilution (e.g. 0.8 litre of seawater + 400 µl of spiking solution) and mix thoroughly. You may use different volumes but the dilution factor should always be kept to a factor of 2000.

Analyse the test materials immediately after preparation. Realize that some determinands can adsorb to the wall of the flask due to limited solubility in seawater of those determinands.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-11	Chlorophyll-a in Seawater
Test materials	QCH086SW and QCH087SW

Objective

This study covers the determination of chlorophyll a, b, c and pheopigments in filtered seawater residue test materials.

Test Materials and storage

The test materials for the analysis of chlorophyll a, b, c and pheopigments were prepared at Wageningen Environmental Research the Netherlands. Test materials were prepared unspiked seawater (QCH086SW) and freshwater (QCH087SW). For each test material, the resultant damp filter paper (Whatman GF/F) was wrapped in aluminium foil, inserted into cryovial and immediately 'flash frozen' in liquid nitrogen. The test materials were stored at -80°C until the day of dispatch. The test materials were homogeneous for the purposes of the LP study.

The filter papers have been shipped on cool packs, and should be stored at -20°C, or a lower temperature, immediately upon receipt, and should be analysed as soon as possible after receipt. The cool packs will thaw slowly during transit, but this does not significantly affect the stability of the material for the purpose of this study, providing the test materials themselves are frozen immediately on receipt.

Sample Code	Description
QCH086SW	Filtered residue from 1 litre of seawater
QCH087SW	Filtered residue from 1 litre of freshwater

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following pigments should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Filtered residues	Const	Prop	
Chlorophyll-a	µg/L	0.1–20	0.05	12.5%	
Chlorophyll-b	µg/L	0.01–5	0.01	12.5%	
Chlorophyll-c	µg/L	0.02–2.5	0.01	12.5%	
Pheopigments	µg/L	0.02–2.5	0.01	12.5%	
Chlorophyll-a (HPLC)	µg/L	0.1–20	0.05	12.5%	
Chlorophyll-b (HPLC)	µg/L	0.01–5	0.01	12.5%	
Chlorophyll-c (HPLC)	µg/L	0.02–2.5	0.01	12.5%	
Chlorophyll-a (corrected)	µg/L	0.1–20	0.05	12.5%	

Chlorophyll-a (corrected) should only be used to report chlorophyll-a concentrations which are corrected for Pheophytin.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

Treat all test materials in the same manner as your routine samples. Use your normal validated methods and procedures to analyse the test materials. Only one result per determinand per test material is required. The results of each determinand should be expressed on the test materials "as received". **Concentrations need to be calculated based on a filter prepared out of a 1 litre sample.**

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site.

Chlorophyll-a (corrected) should only be used to report chlorophyll-a concentrations which are corrected for Phaeophytin.

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

N.B. Laboratories using HPLC related methods should report their results separately as well as labs who correct their results of Chlorophyll-a for Phaeophytine.

Please report your Method information together with the results.

ROUND	2017 - 1
AQ-12	Organotins in Seawater
Test materials	QSP046SW and QSP047SW

Objective

This study covers the determination of organotin compounds in seawater test materials.

Test Materials and storage

QSP046SW has to be spiked in your own laboratory with the spiking solution QSP046SS. QSP046SW is not preserved as it will be spiked in your own laboratory. QSP047SW is preserved with hydrochloric acid and it is ready for analysis.

The test materials QSP046SW and QSP047SW were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK.

It is important that the vial containing the spike solution is used or opened only when you are ready to complete the analysis. Please check the vial to ensure it has not been damaged during transit.

The seawater test materials should be stored in a refrigerator at +4°C, and analyzed as soon as possible after receipt. Once the test materials are opened they should be analyzed immediately.

Sample Code	Description
QSP046SW	Seawater with spiking solution QSP046SS for spiking QSP046SW
QSP047SW	Seawater spiked with organotins

N.B. Please use the correct spiking solution with the correct bottle of seawater. Use the spiking solution labelled QSP046SS for spiking solution ONLY with the bottle of seawater labelled QSP046SW.

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

As a guide, the concentrations of the organotin compounds in the spiked test materials are within the following ranges:

Determinand	Unit	Concentration range	Error		AA-EQS
			Const	Prop	
Dibutyltin(DBT)	ng Sn/kg	1 - 50	0.05	12.5%	0.2
Diphenyltin(DPT)	ng Sn/kg	1 - 100	0.05	12.5%	
Monobutyltin(MBT)	ng Sn/kg	1 - 200	0.05	12.5%	0.2
Monophenyltin(MPT)	ng Sn/kg	1 - 50	0.05	12.5%	
Tributyltin(TBT)	ng Sn/kg	1 - 200	0.05	12.5%	0.2
Triphenyltin(TPT)	ng Sn/kg	1 - 200	0.05	12.5%	

Analysis

A 2000-times dilution of the spiking solutions is required, using the seawater test materials to produce the spiked seawater (E.g. 0.8 litre seawater + 400 µl spiking solution). The dilution procedure is given below:

The spiking solution should be stabilised at room temperature

Weigh an empty flask, which is normally used within your laboratory for this kind of analysis. Weigh the required amount of the seawater provided, in the flask. A mass of 820 gram of water is equal to 800 ml of seawater (density=1.025 kg/L).

Add the spiking solution to the flask with seawater, corresponding to a 2000-times dilution (e.g. 0.8 litre of seawater + 400 µl of spiking solution) and mix thoroughly. You may use different volumes but the dilution factor should always be kept to a factor of 2000.

Analyse the test materials immediately after preparation. Realize that some determinands can adsorb to the wall of the flask due to limited solubility in seawater of those determinands.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site. Only one result per determinand per test material is required.

All results should be reported as **ng Sn/kg seawater**. The concentration of organotins should be determined against your own calibration solutions.

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017-1
AQ-13	Polycyclic Aromatic Hydrocarbons in Seawater
Test materials	QPH028SW, QPH029SW and QPH030SW

Objective

This study covers the determination of PAHs in the seawater test materials

Test Materials and storage

The test materials were prepared at Wageningen Environmental Research, The Netherlands. The seawater used to prepare the test materials was collected from the North Sea between Belgium and the UK. The test materials were prepared in bulk in a 100 litre vessel. The seawater was filtered using a 0.45 µm / 0.2 µm double-membrane filter.

Methanol solutions containing PAHs compounds (QPH028SS spike solution and QPH029SS spike solution) were prepared in bulk and ampouled to make the spiking solutions.

The test material QPH030SW was prepared in bulk in a 100 litre vessel. The seawater was filtered using a 0.45 µm / 0.2 µm double-membrane filter. The low salinity seawater used to prepare test material QPH030SW was spiked with approximately 5 gram highly contaminated (with PAHs), colloidal milled harbour sediment. The vessel with low salinity seawater and fine sediment was mixed intensively for two hours. Following a stagnant period of 30 minutes most of the water layer was pumped into another vessel. Stirring the content of this vessel, flasks were filled with two times 375 g. This sample should be analysed as a so called total water sample.

Test materials should be stored in a refrigerator at +4°C, and analysed as soon as possible after receipt. Once the test materials are opened they should be analysed immediately.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QPH028SW	Seawater with Spiking solution (QPH028SS) (low level: 0.001-0.2 µg/L)
QPH029SW	Seawater with Spiking solution (QPH029SS) (high level: 0.1-10 µg/L)
QPH030SW	Low salinity Seawater spiked using Sediment

Precaution

Some of the substances may presents a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following trace PAHs should be determined and the indicative concentrations are given. These indicative concentrations sometimes differ from the indication ranges given in the Quasimeme guide. In relation to the Seawater spiked with sediment the indicative range is an indication only.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Determinand	Unit	Concentration range		Error		AA-EQS
		Seawater (Sediment Spiked)	Seawater (Spiked)	Const	Prop	
Acenaphthene	µg/L	0.02–200	0.2–200	0.01	12.5%	
Acenaphthylene	µg/L	0.001–10	0.2–200	0.01	12.5%	
Anthracene	µg/L	0.02–200	0.02–20	0.01	12.5%	0.1
Benzo[a]pyrene	µg/L	0.01–100	0.001–1	0.01	12.5%	0.05
Benzo[b]fluoranthene	µg/L	0.01–100	0.001–1	0.01	12.5%	0.03
Benzo[k]fluoranthene	µg/L	0.01–100	0.001–1	0.01	12.5%	0.03
Benzo[g,h,i]perylene	µg/L	0.002–20	0.001–1	0.01	12.5%	0.002
Fluoranthene	µg/L	0.04–400	0.02–20	0.01	12.5%	0.1
Indeno(1,2,3-cd)pyrene	µg/L	0.04–400	0.01–10	0.01	12.5%	0.002
Naphthalene	µg/L	0.01–100	0.2–200	0.01	12.5%	1.2
Phenanthrene	µg/L	0.05–500	0.02–20	0.01	12.5%	

Analysis

Preparation and Analysis

A 2000-times dilution of the spiking solutions is required, using the seawater test materials to produce the spiked seawater (E.g. 0.8 litre seawater + 400 µl spiking solution). The dilution procedure is given below:

The spiking solution should be stabilised at room temperature

Weigh an empty flask, which is normally used within your laboratory for this kind of analysis. Weigh the required amount of the seawater provided, in the flask. A mass of 820 gram of water is equal to 800 ml of seawater (density=1.025 kg/L).

Add the spiking solution to the flask with seawater, corresponding to a 2000-times dilution (e.g. 0.8 litre of seawater + 400 µl of spiking solution) and mix thoroughly. You may use different volumes but the dilution factor should always be kept to a factor of 2000.

Analyse the test materials immediately after preparation. Realize that some determinands can adsorb to the wall of the flask due to limited solubility in seawater of those determinands.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
MS-1	Trace metals in Sediment
Test materials	QTM118MS and QTM119MS

Objective

This study covers the determination of metals, total organic carbon and carbonate in marine sediment test materials.

Test Materials and storage

Test materials were prepared by WEPAL, Wageningen, The Netherlands.

Each batch of material was prepared in bulk. The level of within and between sample homogeneity for the sediment was determined. All materials have been shown to be homogeneous at or below the intake mass used by the participants, and stable for the purposes of the test.

The test materials should be stored at room temperature prior to analysis, and analysed as soon as possible after receipt.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QTM118MS	Sediment (estuarine)
QTM119MS	Sediment (estuarine)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following trace metals should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Sediment	Const	Prop	
Aluminium-AE	%	1–10	0.1	12.5%	
Aluminium-RT	%	1–10	0.1	12.5%	
Arsenic-AE	mg/kg	2–50	1	12.5%	
Arsenic-RT	mg/kg	2–50	1	12.5%	
Cadmium-AE	µg/kg	10–2000	20	12.5%	
Cadmium-RT	µg/kg	10–2000	20	12.5%	
Chromium-AE	mg/kg	10–1000	2	12.5%	
Chromium-RT	mg/kg	10–1000	2	12.5%	
Copper-AE	mg/kg	1–500	1	12.5%	
Copper-RT	mg/kg	1–500	1	12.5%	
Inorganic-carbonate	%	0.05–10	0.05	12.5%	
Iron-AE	%	0.5–10	0.1	12.5%	
Iron-RT	%	0.5–10	0.1	12.5%	
Lead-AE	mg/kg	5–500	2	12.5%	
Lead-RT	mg/kg	5–500	2	12.5%	
Lithium-AE	mg/kg	10–100	0.1	12.5%	
Lithium-RT	mg/kg	10–100	0.1	12.5%	
Manganese-AE	mg/kg	100–2000	0.1	12.5%	
Manganese-RT	mg/kg	100–2000	0.1	12.5%	
Mercury-AE	µg/kg	50–2500	10	12.5%	

Mercury-RT	µg/kg	50—2500	10	12.5%	
Nickel-AE	mg/kg	5—100	1	12.5%	
Nickel-RT	mg/kg	5—100	1	12.5%	
Scandium-AE	mg/kg	1—20	0.1	12.5%	
Scandium-RT	mg/kg	1—20	0.1	12.5%	
TOC	%	0.2—10	0.1	12.5%	
Zinc-AE	mg/kg	20—1500	2.5	12.5%	
Zinc-RT	mg/kg	20—1500	2.5	12.5%	

RT = Real Total destructions e.g. HF-destruction, Röntgen-diffraction and neutron activation
 AE= Acid extractable and all other methods

Total organic carbon and inorganic carbonate can be determined for both test materials. Please note that inorganic carbonate should be reported as % carbon

In addition to the parameters given in the table you will be able to report the following metals: Na, Mg, P, S, K, Ca, Ti, V, Co, Ga, Rb, Se, Sr, Mo, Sn, Cs, Ba, Ce, Ta, Tl, Th, U.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

The sediments have been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator. It is left to the discretion of the laboratory to dry the material. You are advised to check the moisture content of the sediment by drying a small portion of the sediment to constant weight at ~ 110°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material "as received".

The concentration of metals should be determined against your own calibration solutions.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
MS-2 Chlorinated Organics in Sediment	
Test materials	QOR130MS and QOR131MS

Objective

This study covers the determination of chlorobiphenyls (CBs), organochlorine pesticides (OCPs) and total organic carbon in marine sediment test materials.

Test Materials and storage

Test materials were prepared by WEPAL, Wageningen, The Netherlands.

Each batch of material was prepared in bulk. The level of within and between sample homogeneity for the sediment was determined. All materials have been shown to be homogeneous at or below the intake mass used by the participants, and stable for the purposes of the test.

The test materials should be stored at room temperature prior to analysis, and analysed as soon as possible after receipt.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QOR130MS	Sediment (river)
QOR131MS	Sediment (estuarine)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following Chlorinated Organics should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Sediment	Const	Prop	
a-HCH	µg/kg	0.02–1	0.02	12.5%	
b-HCH	µg/kg	0.05–2	0.025	12.5%	
CB101	µg/kg	0.2–50	0.025	12.5%	
CB105	µg/kg	0.1–50	0.025	12.5%	
CB118	µg/kg	0.1–200	0.025	12.5%	
CB138	µg/kg	0.2–50	0.025	12.5%	
CB138+CB163	µg/kg	0.2–50	0.025	12.5%	
CB153	µg/kg	0.2–50	0.025	12.5%	
CB156	µg/kg	0.05–5	0.025	12.5%	
CB180	µg/kg	0.1–50	0.025	12.5%	
CB28	µg/kg	0.1–50	0.025	12.5%	
CB31	µg/kg	0.1–50	0.025	12.5%	
CB52	µg/kg	0.1–500	0.025	12.5%	
d-HCH	µg/kg	0.05–2	0.025	12.5%	
Dieldrin	µg/kg	0.1–10	0.025	12.5%	
g-HCH	µg/kg	0.05–2	0.025	12.5%	
HCB	µg/kg	0.05–250	0.025	12.5%	
HCBD	µg/kg	0.1–10	0.025	12.5%	
op'-DDT	µg/kg	0.02–250	0.025	12.5%	
pp'-DDD	µg/kg	0.1–20	0.025	12.5%	

pp'-DDE	µg/kg	0.1—10	0.025	12.5%	
pp'-DDT	µg/kg	0.1—10	0.025	12.5%	
TOC	%	0.2—10	0.02	12.5%	
Transnonachlor	µg/kg	0.01—2	0.025	12.5%	
PN	%		0.02	12.5%	

In addition total organic carbon should be determined for both test materials.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

The sediments have been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator. It is left to the discretion of the laboratory to dry the material. You are advised to check the moisture content of the sediment by drying a small portion of the sediment to constant weight at ~ 110°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material "as received".

The concentrations should be determined against your own calibration solutions.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
MS-3	Polycyclic Aromatic Hydrocarbons in Sediment
Test materials	QPH093MS and QPH094MS

Objective

This study covers the determination of PAHs and total organic carbon in marine sediment test materials.

Test Materials and storage

Test materials were prepared by WEPAL, Wageningen, The Netherlands.

Each batch of material was prepared in bulk. The level of within and between sample homogeneity for the sediment was determined. All materials have been shown to be homogeneous at or below the intake mass used by the participants, and stable for the purposes of the test.

The test materials should be stored at room temperature prior to analysis, and analysed as soon as possible after receipt.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QPH093MS	Sediment (estuarine)
QPH094MS	Sediment (open sea)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following PAHs and alkylated PAHs should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Sediment	Const	Prop	
1-Methylpyrene	µg/kg	2–500	0.5	12.5%	
2-Methylphenanthrene	µg/kg	5–1000	0.5	12.5%	
3,6-Dimethylphenanthrene	µg/kg	1–500	0.5	12.5%	
Acenaphthene	µg/kg	2–500	0.1	12.5%	
Acenaphthylene	µg/kg	1–100	0.2	12.5%	
Anthracene	µg/kg	2–500	0.1	12.5%	
Benzo[a]anthracene	µg/kg	10–1500	0.1	12.5%	
Benzo[a]fluorene	µg/kg	10–1000	0.5	12.5%	
Benzo[a]pyrene	µg/kg	10–1500	0.1	12.5%	
Benzo[b]fluoranthene	µg/kg	10–1500	0.5	12.5%	
Benzo[e]pyrene	µg/kg	10–1500	0.2	12.5%	
Benzo[g,h,i]perylene	µg/kg	10–1500	0.2	12.5%	
Benzo[k]fluoranthene	µg/kg	10–1000	0.1	12.5%	
Benzo[fluoranthenes (a+b+j+k)	µg/kg	10–2000	0.1	12.5%	
Chrysene	µg/kg	10–1500	0.2	12.5%	
Chrysene+Triphenylene	µg/kg	10–3000	0.2	12.5%	
Dibenz[a,h]anthracene	µg/kg	5–500	0.05	12.5%	
Dibenzo[a,i]pyrene	µg/kg		0.5	12.5%	
Dibenzothiophene	µg/kg	2–200	0.1	12.5%	
Fluoranthene	µg/kg	20–3000	0.2	12.5%	

Fluorene	µg/kg	2—300	0.1	12.5%	
Indeno[1,2,3-cd]pyrene	µg/kg	10—1500	0.2	12.5%	
Naphthalene	µg/kg	10—1500	0.5	12.5%	
Perylene	µg/kg	10—500	0.2	12.5%	
Phenanthrene	µg/kg	10—2000	0.5	12.5%	
Pyrene	µg/kg	10—3000	0.2	12.5%	
TOC	%	0.2—10	0.02	12.5%	
Triphenylene	µg/kg	20—3000	0.5	12.5%	
C1-phenanthrenes/anthracenes	µg/kg		0.5	12.5%	
C2-phenanthrenes/anthracenes	µg/kg		0.5	12.5%	
C3-phenanthrenes/anthracenes	µg/kg		0.5	12.5%	
C1-pyrenes/fluoranthenes	µg/kg		0.5	12.5%	
C2-pyrenes/fluoranthenes	µg/kg		0.5	12.5%	
C1-chrysenes	µg/kg		0.5	12.5%	
C2-chrysenes	µg/kg		0.5	12.5%	
C1-benzofluoranthenes	µg/kg		0.5	12.5%	
PN	%		0.02	12.5%	

In addition total organic carbon should be determined for both test materials.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

The sediments have been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator. It is left to the discretion of the laboratory to dry the material. You are advised to check the moisture content of the sediment by drying a small portion of the sediment to constant weight at ~ 110°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. You may use any method with the appropriate extraction and clean-up. The final determination may be made using GC, GC-MS, HPLC etc. The method codes should be entered fully on the exercise template.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material “as received”.

The concentrations should be determined against your own calibration solutions.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
MS-6	Organotins in Sediment
Test materials	QSP060MS and QSP061MS

Objective

This study covers the determination of organotin compounds in sediment test materials.

Test Materials and storage

The sediment test materials were supplied by WEPAL, Wageningen.

The dry sediment test materials should be stored at room temperature, in a dry place, prior to analysis, and analysed as soon as possible after receipt.

Each batch of material was prepared in bulk. The level of within and between sample homogeneity for the sediment was determined. All materials have been shown to be homogeneous at or below the intake mass used by the participants, and stable for the purposes of the test.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QSP060MS	Sediment (estuarine)
QSP061MS	Sediment (estuarine)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following organotin compounds should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Sediment	Const	Prop	
Dibutyltin(DBT)	µg Sn/kg	1—500	0.1	12.5%	
Diphenyltin(DPT)	µg Sn/kg	0.1—200	0.1	12.5%	
Monobutyltin(MBT)	µg Sn/kg	1—500	0.1	12.5%	
Monophenyltin(MPT)	µg Sn/kg	0.1—200	0.1	12.5%	
Tributyltin(TBT)	µg Sn/kg	1—500	0.1	12.5%	
Triphenyltin(TPT)	µg Sn/kg	0.1—200	0.1	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

The sediments have been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. Report your method codes using the Data Submission Form on the Participant Site. Please check each of your method codes and update where necessary. Advise QUASIMEME of additional codes that would better describe your methodology.

Only one result per determinand per test material is required.

The results of each determinand should be expressed as Sn on the test materials "as received". All results should be reported as $\mu\text{g Sn /kg}$ weight of sediment as received

The concentration of organotins should be determined against your own calibration solutions.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
MS-7 Brominated flame retardants in Sediment	
Test materials	QBC050MS and QBC051MS

Objective

This study covers the determination of brominated compounds in sediment test material.

Test Materials and storage

The test materials were supplied by WEPAL, Wageningen, The Netherlands.

The Sediment has been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator. It is left to the discretion of the laboratory to dry the material. You are advised to check the moisture content of the sediment by drying a small portion of the sediment to constant weight at ~ 110°C.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored in a dry place at room temperature in the dark, prior to analysis, and analysed as soon as possible after receipt.

Sample Code	Description
QBC050MS	Sediment (estuarine)
QBC051MS	Sediment (open sea)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Sediment	Const	Prop	
a-HBCD	µg/kg		0.05	12.5%	
BDE100	µg/kg	0.01–10	0.05	12.5%	
BDE153	µg/kg	0.1–5	0.05	12.5%	
BDE154	µg/kg	0.01–5	0.05	12.5%	
BDE183	µg/kg	0.1–2	0.05	12.5%	
BDE209	µg/kg	20–200	0.05	12.5%	
BDE28	µg/kg	0.01–2	0.05	12.5%	
BDE47	µg/kg	0.1–20	0.05	12.5%	
BDE66	µg/kg	0.01–10	0.05	12.5%	
BDE85	µg/kg	0.01–10	0.05	12.5%	
BDE99	µg/kg	0.1–50	0.05	12.5%	
b-HBCD	µg/kg		0.05	12.5%	
Dimethyl-TBBP-A	µg/kg		0.05	12.5%	
g-HBCD	µg/kg		0.05	12.5%	
TBBP-A	µg/kg		0.05	12.5%	
Total-HBCD	µg/kg	50–200	0.05	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

The sediments have been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. Report your method codes using the Data Submission Form on the Participant Site. Please check each of your method codes and update where necessary. Advise QUASIMEME of additional codes that would better describe your methodology.

Only one result per determinand per test material is required.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
MS-8	PFASs in Sediment
Test materials	QPF001MS and QPF002MS

Objective

This study covers the determination of perfluorinated alkyl substances (PFASs) in sediment test material.

Test Materials and storage

The test materials were supplied by WEPAL, Wageningen, The Netherlands.

The Sediment has been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator. It is left to the discretion of the laboratory to dry the material. You are advised to check the moisture content of the sediment by drying a small portion of the sediment to constant weight at ~ 110°C.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored in a dry place at room temperature in the dark, prior to analysis, and analysed as soon as possible after receipt.

Sample Code	Description
QPF001MS	Sediment (harbour)
QPF002MS	Sediment (river)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Sediment	Const	Prop	
PFOS	µg/kg		0.005	12.5%	
PFBA	µg/kg		0.005	12.5%	
PFPeA	µg/kg		0.005	12.5%	
PFHxA	µg/kg		0.005	12.5%	
PFHpA	µg/kg		0.005	12.5%	
PFOA	µg/kg		0.005	12.5%	
PFNA	µg/kg		0.005	12.5%	
PFUnA	µg/kg		0.005	12.5%	
PFDoA	µg/kg		0.005	12.5%	
PFTTrA	µg/kg		0.005	12.5%	
PFTeA	µg/kg		0.005	12.5%	
L-PFBS	µg/kg		0.005	12.5%	
L-PFHxS	µg/kg		0.005	12.5%	
L-PFHps	µg/kg		0.005	12.5%	
PFOSA	µg/kg		0.005	12.5%	
PFDS	µg/kg		0.005	12.5%	
PFODA	µg/kg		0.005	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

The sediments have been dried and sieved to < 0.5 mm. Before sub-sampling, shake the bottle for at least two minutes. Repeat this re-homogenisation each time the test material is used. It is recommended that the sediments be kept in a desiccator.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. Report your method codes using the Data Submission Form on the Participant Site. Please check each of your method codes and update where necessary. Advise QUASIMEME of additional codes that would better describe your methodology.

Only one result per determinand per test material is required.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
BT-1	Trace metals in Biota
Test materials	QTM114BT and QTM115BT

Objective

This study covers the determination of ten trace metals, ash weight, dry weight and total lipid in biological tissue test materials.

Test Materials and storage

The test materials were supplied by Wageningen Marine Research (IMARES), IJmuiden, The Netherlands.

The tins or jars contains approximately 70g of minced sterilised material, to which butylhydroxytoluene (BHT) has been added as an antioxidant. Representative sub samples were dispensed into tins and sterilised by autoclaving. Each batch of material was prepared in bulk.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored at room temperature, prior to analysis, and analysed as soon as possible after receipt.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QTM114BT	Mussel tissue
QTM115BT	Fish fillet (0-5% fat)

Precaution

The jars with biological tissue test materials are filled almost to the brim, so should be opened carefully so that no tissue or moisture is spilt. The jars are filled in this way in order to eliminate as much air as possible. This minimises any degradation of the test materials in transit and storage prior to opening. The jars of biota should be cooled for a few hours at ca. 4°C prior to opening, to eliminate possible overpressure.

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following trace metals should be determined:

Determinand	Unit	Concentration range		Error		AA-EQS
		Shellfish tissue	Fish muscle tissue	Const	Prop	
Arsenic	mg/kg	0.2–10	1–10	0.02	12.5%	
Ash-weight	%			0.1	12.5%	
Cadmium	µg/kg	10–500	1–50	20	12.5%	
Chromium	µg/kg	10–5000	50–500	20	12.5%	
Copper	µg/kg	50–10000	100–1000	100	12.5%	
Dry-weight	%			0.1	12.5%	
Extractable-Lipid	%			0.1	12.5%	
Lead	µg/kg	10–1000	10–50	5	12.5%	
Mercury	µg/kg	5–500	20–1000	20	12.5%	
Nickel	µg/kg	10–2000	10–200	20	12.5%	

Selenium	µg/kg	200—1000	200—2000	10	12.5%	
Silver	µg/kg	1—500	0.5—50	5	12.5%	
Total-Lipid	%			0.1	12.5%	
Zinc	mg/kg	2—200	2—10	2	12.5%	

Ash weight, dry weight and total lipid should also be determined. In addition to the parameters given in the table, also the following metals can be reported: Li, Be, Na, Mg, Al, P, S, K, Ca, Sc, Ti, V, Mn, Fe, Co, Ga, Rb, Sr, Y, Zr, Mo, Pd, Sn, Sb, Te, Cs, Ba, La, Ce, Nd, Ta, W, Pt, Au, Tl, Bi, Th, U. If you normally measure extractable lipid, there is a field in the template for reporting this measurement. This has been added following the request from a number of participants. However, we would encourage you to also report total lipid. A successful QUASH study has found no significant difference between the Smedes lipid method and the Bligh and Dyer method for total lipid determination. We therefore recommend that wherever possible you use the Smedes lipid method, as it gives better reproducibility and does not involve the use of chlorinated solvents. A copy of the Smedes lipid method protocol is included in Annex 3.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

During the sterilisation process moisture may have partly separated from the solid part of the test materials, and this may cause inhomogeneity. Therefore, prior to taking an aliquot for analysis the following treatment of the test materials is essential. After opening the test material container the complete contents should be thoroughly homogenised. As it is difficult to homogenise the sample in the tin, we recommend transferring the material to a larger container for homogenisation and sub-sampling. Aliquots should be taken immediately after homogenising, to prevent liquid lipids from separating from the solid tissue particles, causing inhomogeneity again. After sub-sampling the remaining tissue should be immediately transferred to a clean glass jar, sealed and stored at -20°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

The concentrations should be determined against your own calibration solutions.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material "as received" i.e. on a wet weight basis. If your normal method is to dry the test material prior to extraction then the dry weight must be determined. The final result should be recalculated and reported based on wet weight using the dry weight determination.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
BT-2 Chlorinated Organics in Biota	
Test materials	QOR130BT and QOR131BT

Objective

This study covers the determination chlorobiphenyls (CBs), organochlorine pesticides (OCPs) and total lipid in biological tissue test materials.

Test Materials and storage

The test materials were supplied by Wageningen Marine Research (IMARES), IJmuiden, The Netherlands.

The tins or jars contains approximately 70g of minced sterilised material, to which butylhydroxytoluene (BHT) has been added as an antioxidant. Representative sub samples were dispensed into tins and sterilised by autoclaving. Each batch of material was prepared in bulk.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored at room temperature, in a dry place, prior to analysis, and analysed as soon as possible after receipt. Although the tins have a protective coating, moisture can cause corrosion of the surface of the tins. Apart from the moisture, temperatures below 0°C will not negatively affect the quality in the material, and we do not expect the contaminants to be affected at these temperatures. Temperatures above 30°C may be disadvantageous for the material. The test materials should be stored at room temperature, prior to analysis, and analysed as soon as possible after receipt.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QOR130BT	Whole fish tissue (0-5% fat)
QOR131BT	Mussel tissue

Shellfish tissues might be spiked with organochlorines, so patterns measured can differ from natural patterns

Precaution

The tins with biological tissue test materials are filled almost to the brim, so should be opened carefully so that no tissue or moisture is spilt. The tins are filled in this way in order to eliminate as much air as possible. This minimises any degradation of the test materials in transit and storage prior to opening. The tins of biota should be cooled for a few hours at ca. 4°C prior to opening, to eliminate possible overpressure.

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range			Error		AA-EQS
		Fish Liver tissue and Freshwater Fish	Fish Muscle Tissue	Shellfish Tissue	Const	Prop	
PCB28	µg/kg	1–50	0.05–5	0.05–5	0.025	12.5%	
PCB31	µg/kg	1–10	0.03–3	0.03–3	0.025	12.5%	
PCB52	µg/kg	10–100	0.05–5	0.05–5	0.025	12.5%	
PCB101	µg/kg	30–300	0.1–20	0.1–20	0.025	12.5%	
PCB105	µg/kg	2–100	0.05–10	0.05–10	0.025	12.5%	
PCB118	µg/kg	20–300	0.2–20	0.2–20	0.025	12.5%	
PCB138+PCB163	µg/kg	20–600	0.3–30	0.3–30	0.025	12.5%	
PCB138	µg/kg	20–600	0.3–30	0.3–30	0.025	12.5%	
PCB153	µg/kg	50–1000	0.4–40	0.4–40	0.025	12.5%	
PCB156	µg/kg	1–40	0.03–10	0.03–10	0.025	12.5%	
PCB180	µg/kg	10–200	0.05–5	0.05–5	0.025	12.5%	
□-HCH	µg/kg	0.05–5	0.05–5	0.05–5	0.02	12.5%	
□-HCH	µg/kg	0.1–5	0.05–5	0.05–5	0.025	12.5%	
□-HCH	µg/kg	0.05–5	0.05–5	0.05–5	0.025	12.5%	
□-HCH	µg/kg	0.05–5	0.05–5	0.05–5	0.025	12.5%	
HCB	µg/kg	2–50	0.02–5	0.02–5	0.025	12.5%	
HCBD	µg/kg	0.05–5			0.025	12.5%	
Dieldrin	µg/kg	0.5–100	0.2–20	0.2–20	0.025	12.5%	
pp'-DDD	µg/kg	5–100	0.1–10	0.1–10	0.025	12.5%	
pp'-DDE	µg/kg	10–500	0.3–30	0.3–30	0.025	12.5%	
op'-DDT	µg/kg	0.1–2	0.01–1	0.01–1	0.025	12.5%	
pp'-DDT	µg/kg	0.1–10	0.1–10	0.1–10	0.025	12.5%	
Transnonachlor	µg/kg	0.2–40	0.02–10	0.02–10	0.025	12.5%	
Total-Lipid	%				0.1	12.5%	
Extractable-Lipid	%				0.1	12.5%	

In addition total lipid should be determined. If you normally measure extractable lipid, there is a field in the template for reporting this measurement. This has been added following the request from a number of participants. However, we would encourage you to also report total lipid. A successful QUASH study has found no significant difference between the Smedes lipid method and the Bligh and Dyer method for total lipid determination. We therefore recommend that wherever possible you use the Smedes lipid method, as it gives better reproducibility and does not involve the use of chlorinated solvents. A copy of the Smedes lipid method protocol is included in Annex 3.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

During the sterilisation process moisture may have partly separated from the solid part of the test materials, and this may cause inhomogeneity. Therefore, prior to taking an aliquot for analysis the following treatment of the test materials is essential. After opening the test material container the complete contents should be thoroughly homogenised. As it is difficult to homogenise the sample in the tin, we recommend transferring the material to a larger container for homogenisation and sub-sampling. Aliquots should be taken immediately after homogenising, to prevent liquid lipids from separating from the solid tissue particles, causing inhomogeneity again. After sub-sampling the remaining tissue should be immediately transferred to a clean glass jar, sealed and stored at -20°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

The concentrations should be determined against your own calibration solutions.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material "as received" i.e. on a wet weight basis. If your normal method is to dry the test material prior to extraction then the dry weight must be determined. The final result should be recalculated and reported based on wet weight using the dry weight determination.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
BT-4	Polycyclic Aromatic Hydrocarbons in Biota
Test materials	QPH085BT and QPH086BT

Objective

This study covers the determination of PAHs and total lipid in biological tissue test materials.

Test Materials and storage

The test materials were supplied by Wageningen Marine Research (IMARES), IJmuiden, The Netherlands and the Institute for Environmental Studies, Free University, Amsterdam, The Netherlands.

The tins or jars contains approximately 70g of minced sterilised material, to which butylhydroxytoluene (BHT) has been added as an antioxidant. Representative sub samples were dispensed into tins and sterilised by autoclaving. Each batch of material was prepared in bulk.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored at room temperature, prior to analysis, and analysed as soon as possible after receipt.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QPH085BT	Shellfish tissue (mussel)
QPH086BT	Shellfish tissue (mussel)

Shellfish tissues might be spiked with PAHs, so patterns measured can differ from natural patterns

Precaution

The tins with biological tissue are filled almost to the brim, so should be opened carefully so that no tissue or moisture is spilt. The tins are filled in this way in order to eliminate as much air as possible. This minimizes any degradation of the test materials in transit and storage prior to opening. The tins of biota should be cooled for a few hours at ca. 4°C prior to opening, to eliminate possible overpressure. Opening the bottom of the tin, instead of the top, may also help to prevent spillage of moisture.

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following PAHs should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Shellfish tissue	Const	Prop	
1-Methylpyrene	µg/kg		2	12.5%	
2-Methylphenanthrene	µg/kg	0.2–5	2	12.5%	
3,6-Dimethylphenanthrene	µg/kg	0.2–2	0.5	12.5%	
Acenaphthene	µg/kg	0.5–100	0.2	12.5%	
Acenaphthylene	µg/kg	0.2–5	0.2	12.5%	
Anthracene	µg/kg	0.2–10	0.2	12.5%	
Benzo[a]anthracene	µg/kg	0.2–10	0.2	12.5%	

Benzo[a]fluorene	µg/kg		0.5	12.5%	
Benzo[a]pyrene	µg/kg		0.2–5	0.2	12.5%
Benzo[b]fluoranthene	µg/kg		0.2–10	0.2	12.5%
Benzo[e]pyrene	µg/kg		0.2–10	0.2	12.5%
Benzo[g,h,i]perylene	µg/kg		0.2–5	0.2	12.5%
Benzo[k]fluoranthene	µg/kg		0.2–5	0.2	12.5%
Benzo[fluoranthenes (a+b+j+k)	µg/kg		0.2–20	0.2	12.5%
Chrysene	µg/kg		0.2–20	0.2	12.5%
Chrysene+Triphenylene	µg/kg		0.2–20	0.2	12.5%
Dibenzo[a,h]anthracene	µg/kg		0.2–2	0.1	12.5%
Dibenzo[a,i]pyrene	µg/kg			0.5	12.5%
Dibenzothiophene	µg/kg		0.2–5	0.5	12.5%
Extractable-Lipid	%			0.1	12.5%
Fluoranthene	µg/kg		5–50	0.2	12.5%
Fluorene	µg/kg		1–50	0.2	12.5%
Indeno[1,2,3-cd]pyrene	µg/kg		0.2–5	0.2	12.5%
Naphthalene	µg/kg		1–100	0.2	12.5%
Perylene	µg/kg		0.1–5	0.5	12.5%
Phenanthrene	µg/kg		2–50	0.2	12.5%
Pyrene	µg/kg		1–20	0.2	12.5%
Total-Lipid	%			0.1	12.5%
Triphenylene	µg/kg			5	12.5%

In addition total lipid should be determined. If you normally measure extractable lipid, there is a field in the template for reporting this measurement. This has been added following the request from a number of participants. However, we would encourage you to also report total lipid. A successful QUASH study has found no significant difference between the Smedes lipid method and the Bligh and Dyer method for total lipid determination. We therefore recommend that wherever possible you use the Smedes lipid method, as it gives better reproducibility and does not involve the use of chlorinated solvents. A copy of the Smedes lipid method protocol is included in Annex 3.

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

During the sterilisation process moisture may have partly separated from the solid part of the test materials, and this may cause inhomogeneity. Therefore, prior to taking an aliquot for analysis the following treatment of the test materials is essential. After opening the test material container the complete contents should be thoroughly homogenised. As it is difficult to homogenise the sample in the tin, we recommend transferring the material to a larger container for homogenisation and sub-sampling. Aliquots should be taken immediately after homogenising, to prevent liquid lipids from separating from the solid tissue particles, causing inhomogeneity again. After sub-sampling the remaining tissue should be immediately transferred to a clean glass jar, sealed and stored at -20°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

The concentrations should be determined against your own calibration solutions.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material “as received” i.e. on a wet weight basis. If your normal method is to dry the test material prior to extraction then the dry weight must be determined. The final result should be recalculated and reported based on wet weight using the dry weight determination.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
BT-8	Organotins in Biota
Test materials	QSP060BT and QSP061BT

Objective

This study covers the determination of organotin compounds in biological tissue test materials.

Test Materials and storage

The test materials were supplied by Wageningen Marine Research (IMARES), IJmuiden, The Netherlands and the Institute for Environmental Studies, Free University, Amsterdam, The Netherlands.

The tins or jars contains approximately 70g of minced sterilised material, to which butylhydroxytoluene (BHT) has been added as an antioxidant. Representative sub samples were dispensed into tins and sterilised by autoclaving. Each batch of material was prepared in bulk.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored at room temperature, prior to analysis, and analysed as soon as possible after receipt.

Treat all test materials in the same manner as your routine samples.

Sample Code	Description
QSP060BT	Shrimp tissue
QSP061BT	Shellfish tissue (mussel)

Shellfish tissues might be spiked with organotins, so patterns measured can differ from natural patterns

Precaution

The tins or jars with biological tissue are filled almost to the brim, so should be opened carefully so that no tissue or moisture is spilt. The tins or jars are filled in this way in order to eliminate as much air as possible. This minimizes any degradation of the test materials in transit and storage prior to opening. The tins or jars of biota should be cooled for a few hours at ca. 4°C prior to opening, to eliminate possible overpressure.

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Biota	Const	Prop	
Dibutyltin(DBT)	µg Sn/kg	1–100	0.1	12.5%	
Diphenyltin(DPT)	µg Sn/kg		0.1	12.5%	
Monobutyltin(MBT)	µg Sn/kg	5–30	0.1	12.5%	
Monophenyltin(MPT)	µg Sn/kg		0.1	12.5%	
Tributyltin(TBT)	µg Sn/kg	2–50	0.1	12.5%	
Triphenyltin(TPT)	µg Sn/kg		0.1	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

During the sterilisation process moisture may have partly separated from the solid part of the test materials, and this may cause inhomogeneity. Therefore, prior to taking an aliquot for analysis the following treatment of the test materials is essential. After opening the test material container the complete contents should be thoroughly homogenised. As it is difficult to homogenise the sample in the tin, we recommend transferring the material to a larger container for homogenisation and sub-sampling. Aliquots should be taken immediately after homogenising, to prevent liquid lipids from separating from the solid tissue particles, causing inhomogeneity again. After sub-sampling the remaining tissue should be immediately transferred to a clean glass jar, sealed and stored at -20°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

The concentrations should be determined against your own calibration solutions.

Only one result per determinand per test material is required. The results of each determinand should be expressed as Sn on the test material "as received" i.e. on a wet weight basis. If your normal method is to dry the test material prior to extraction then the dry weight must be determined. The final result should be recalculated and reported based on wet weight using the dry weight determination.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Form which is placed on the Participant Site. It is not possible to report two sets of data using different methods on the same Data Submission Form.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form

Please report your Method information together with the results.

ROUND	2017 - 1
BT-9	Brominated Flame Retardants in Biota
Test materials	QBC050BT and QBC051BT

Objective

This study covers the determination of brominated compounds in biota.

Test Materials and storage

The test materials were supplied by Wageningen Marine Research (IMARES), IJmuiden, The Netherlands and the Institute for Environmental Studies, Free University, Amsterdam, The Netherlands.

The tins or jars contains approximately 70g of minced sterilised material, to which butylhydroxytoluene (BHT) has been added as an antioxidant. Representative sub samples were dispensed into tins and sterilised by autoclaving. Each batch of material was prepared in bulk.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored at room temperature, prior to analysis, and analysed as soon as possible after receipt.

Sample Code	Description
QBC050BT	Turbot liver tissue
QBC051BT	Freswater fish tissue

Shellfish tissues might be spiked with BFRs, so patterns measured can differ from natural patterns

Precaution

The tins or jars with biological tissue are filled almost to the brim, so should be opened carefully so that no tissue or moisture is spilt. The tins or jars are filled in this way in order to eliminate as much air as possible. This minimises any degradation of the test materials in transit and storage prior to opening. The tins or jars of biota should be cooled for a few hours at ca. 4°C prior to opening, to eliminate possible overpressure.

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Biota	Const	Prop	
BDE28	µg/kg	0.001—1	0.005	12.5%	
BDE47	µg/kg	0.05—40	0.005	12.5%	
BDE66	µg/kg	0.01—10	0.005	12.5%	
BDE85	µg/kg	0.01—10	0.005	12.5%	

BDE99	µg/kg	0.01—10	0.005	12.5%	
BDE100	µg/kg	0.005—10	0.005	12.5%	
BDE153	µg/kg	0.01—2	0.005	12.5%	
BDE154	µg/kg	0.001—5	0.005	12.5%	
BDE183	µg/kg	0.001—0.1	0.005	12.5%	
BDE209	µg/kg	0.01—0.1	0.005	12.5%	
TBBP-A	µg/kg		0.005	12.5%	
Dimethyl-TBBP-A	µg/kg		0.005	12.5%	
□-HBCD	µg/kg		0.005	12.5%	
□-HBCD	µg/kg		0.005	12.5%	
□-HBCD	µg/kg		0.005	12.5%	
Total-HBCD	µg/kg		0.005	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

During the sterilisation process of the biological tissue test material moisture may have partly separated from the solid part of the test materials, and this may cause inhomogeneity. Therefore, prior to taking an aliquot for analysis the following treatment of the test materials is essential. After opening the test material container the complete contents should be thoroughly homogenised. As it is difficult to homogenise the sample in the jar, we recommend transferring the material to a larger container for homogenisation and sub-sampling. Aliquots should be taken immediately after homogenising, to prevent liquid lipids from separating from the solid tissue particles, causing inhomogeneity again. After sub-sampling the remaining tissue should be immediately transferred to a clean glass jar, sealed and stored at -20°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. Recovery values of over 100% should not be used to correct the data.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material "as received" i.e. on a wet weight basis. If your normal method is to dry the test material prior to extraction then the dry weight must be determined. The final result should be recalculated and reported based on wet weight using the dry weight determination.

You may wish to use two different GC columns of different polarity for the determination of the BDEs. Use your own judgement to report the best result. The column used for this result should be reported.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site.

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
BT-10	Perfluorinated Alkyl Substances (PFASs) in Biota
Test materials	QPF005BT and QPF006BT

Objective

This study covers the determination of perfluorinated alkyl substances in biota.

Test Materials and storage

The test materials were supplied by Wageningen Marine Research (IMARES), IJmuiden, The Netherlands and the Institute for Environmental Studies, Free University, Amsterdam, The Netherlands.

The tins or jars contains approximately 70g of minced sterilised material, to which butylhydroxytoluene (BHT) has been added as an antioxidant. Representative sub samples were dispensed into tins and sterilised by autoclaving. Each batch of material was prepared in bulk.

All materials have been shown to be homogeneous at or below the intake mass used by the participants, and are stable for the purposes of the test.

The test materials should be stored at room temperature, prior to analysis, and analysed as soon as possible after receipt.

Sample Code	Description
QPF005BT	Turbot liver tissue
QPF006BT	Flounder whole fish tissue

Precaution

The tins or jars with biological tissue are filled almost to the brim, so should be opened carefully so that no tissue or moisture is spilt. The tins or jars are filled in this way in order to eliminate as much air as possible. This minimises any degradation of the test materials in transit and storage prior to opening. The tins or jars of biota should be cooled for a few hours at ca. 4°C prior to opening, to eliminate possible overpressure.

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The following analytes should be determined:

Determinand	Unit	Concentration range	Error		AA-EQS
		Biota	Const	Prop	
PFOS	ng/kg	0.1 - 1000	0.05	12.5%	
PFBA	ng/kg		0.05	12.5%	
PFPeA	ng/kg		0.05	12.5%	
PFHxA	ng/kg		0.05	12.5%	
PFHpA	ng/kg		0.05	12.5%	
PFOA	ng/kg		0.05	12.5%	
PFNA	ng/kg		0.05	12.5%	

PFDA	ng/kg	0.1 - 10	0.05	12.5%	
PFUdA	ng/kg	0.001 - 1	0.05	12.5%	
PFDaA	ng/kg	0.001 - 0.1	0.05	12.5%	
PFTTrDA	ng/kg	0.01 - 0.1	0.05	12.5%	
PFTeDA	ng/kg		0.05	12.5%	
L-PFBS**	ng/kg	0.1 - 10	0.05	12.5%	
L-PFHxS**	ng/kg		0.05	12.5%	
L-PFHps**	ng/kg		0.05	12.5%	
PFOSA	ng/kg	0.1 - 10	0.05	12.5%	
PFDS	ng/kg		0.05	12.5%	
PFODA	ng/kg		0.05	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

During the sterilisation process of the biological tissue test material moisture may have partly separated from the solid part of the test materials, and this may cause inhomogeneity. Therefore, prior to taking an aliquot for analysis the following treatment of the test materials is essential. After opening the test material container the complete contents should be thoroughly homogenised. As it is difficult to homogenise the sample in the jar, we recommend transferring the material to a larger container for homogenisation and sub-sampling. Aliquots should be taken immediately after homogenising, to prevent liquid lipids from separating from the solid tissue particles, causing inhomogeneity again. After sub-sampling the remaining tissue should be immediately transferred to a clean glass jar, sealed and stored at -20°C.

Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. Recovery values of over 100% should not be used to correct the data.

Only one result per determinand per test material is required. The results of each determinand should be expressed on the test material "as received" i.e. on a wet weight basis. If your normal method is to dry the test material prior to extraction then the dry weight must be determined. The final result should be recalculated and reported based on wet weight using the dry weight determination.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site.

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
BT-7	ASP Shellfish Toxins
Test materials	QST222SS, QST223BT and QST224BT

Objective

This study covers the determination of amnesic shellfish toxins domoic acid and epidomoic acid (as a racemic mixture) in standard solution and shellfish tissue test materials.

Test Materials and storage

The test materials were supplied by the Marine Institute, Galway, Ireland.

- QST222SS is a ASP standard solution.
- QST223BT is a clam tissue homogenate supplied in a plastic vial.
- QST224BT is an scallop tissue homogenate supplied in a plastic vial.

For QST223BT and QST224BT, each vial contains sufficient material for one-shot analysis of domoic and epidomoic acid.

Each batch of material was prepared in bulk. The level of within and between sample homogeneity and stability was determined. All materials have been shown to be homogeneous at or below the intake mass normally used, and stable for the purposes of the test.

Begin the analysis as soon as possible, preferably within 7 days of receipt.

The shellfish tissue homogenates (5ml plastic vials) should be stored at -20°C, or a lower temperature, immediately upon receipt, until analysis.

The standard solution QST222SS should be stored in the refrigerator at ca 4°C immediately upon receipt, until analysis

The test materials have been shipped on cool packs. The cool packs will thaw slowly during transit, but this does not significantly affect the stability of the material for the purpose of this study, providing the test materials themselves are stored as directed, immediately on receipt.

Sample Code	Description
QST222SS	Standard Solution
QST223BT	Shellfish tissue (clam tissue)
QST224BT	Shellfish tissue (scallop tissue)

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

Report the sum of the domoic acid and epidomoic acid as a racemic mixture.

Determinand	Unit	Concentration range	Error		AA-EQS
		Shellfish tissue	Const	Prop	
Domoic+Epidomoic	mg/kg		0.1	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Analysis

It is advisable to analyse the test materials as soon as possible after receipt. Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. One result per test material is required, for the sum of domoic and epidomoic acid as a racemic mixture. The results of each determinand should be expressed on the test material "as received" i.e. on a wet weight basis. The concentrations should be determined against your own calibration solutions.

Each vial contains sufficient quantity of homogenate for one analysis. The whole transferable contents of each vial should be extracted, and one result reported for the sum of the two isomers. To transfer the contents into a preweighed or tared extraction tube, the vial should be fully defrosted, vortex-mixed and the contents poured into the desired container.

All results should be reported in mg/kg on the basis of wet weight of the test material as provided. The density of the standard solution is 0.9853 g/ml at 22°C. The weight of shellfish tissue test materials should be determined prior to analysis.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
BT-11	Lipophilic Shellfish Toxins
Test materials	QST225SS, QST226SS, QST227BT, QST228BT and QST229BT

Objective

This study covers the determination of lipophilic toxins in shellfish tissue.

Test Materials and storage

The test materials were supplied by the Marine Institute, Galway, Ireland.

Each vial contains sufficient material for one-shot analysis of OA, AZA, YTX and PTX-group toxins..

Each batch of material was prepared in bulk. The level of within and between sample homogeneity and stability was determined. All materials have been shown to be homogeneous at the intake mass normally used, and stable for the purposes of the test.

Begin the analysis as soon as possible, preferably within 7 days of receipt.

All materials (contained either in ampoules or in 5ml plastic vials) should be stored at -20°C, or a lower temperature, immediately upon receipt, until analysis.

The test materials have been shipped on cool packs. The cool packs will thaw slowly during transit, but this does not significantly affect the stability of the material for the purpose of this study, providing the test materials themselves are stored as directed, immediately on receipt.

Sample Code	Description
QST225SS	AZA standard solution
QST226SS	Lipophilic standard solution
QST227BT	DSP-AZP extract
QST228BT	Mussel tissue homogenate
QST229BT	Mussel tissue homogenate

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

Determinands

a) Methods based on chromatographic separation techniques (e.g. LC-FD, or LC-MS):

Please report lipophilic toxins (if detected) as follows:

- Free OA-group toxins, OA, DTX-1, DTX-2 (pre-hydrolysis) individually and their sum (standard solutions, extracts and tissues).
- Total OA-group toxins, OA, DTX-1, DTX-2 (post-hydrolysis) individually and their sum (extracts and tissues only).
- PTX-group toxins, PTX-1 and PTX-2 individually (standard solutions, extracts and tissues).
- Total OA-group and PTX-group toxins, sum of OA, DTX-1, DTX-2 (post-hydrolysis), PTX-1 and PTX-2 (extracts and tissues only).
- AZA-group toxins, AZA-1, AZA-2 and AZA-3 individually and their sum (standard solutions, extracts and tissues).

- YTX-group toxins, YTX, homo-YTX, 45-OH-YTX and 45-OH-homo-YTX individually and their sum (standard solutions, extracts and tissues).

Please note, for the OA-group toxins, there is no result reported for the ester-forms themselves, only for free toxins and the sum of free toxins plus esters.

b) Methods based on determination of the sum of OA-equivalents present (e.g. PP2a):

For the standard solution, report the sum of OA-equivalents as free toxins (without hydrolysis), and the sum of OA-equivalents post hydrolysis. This means there is no result reported for the esters-forms themselves, only for free toxins and the sum of free toxins plus esters.

If you do not analyse for one of the determinands, eg. DTX-1 or DTX-2, please do not report the sum of OA+DTX-1+DTX-2. Equally if you do not carry out hydrolysis or determination of DTX-1 or DTX-2 post-hydrolysis, please do not report the sum of hydrolysed results.

Determinand	Unit	Concentration range	Error		AA-EQS
			Const	Prop	
AZA-1	µg/kg		0.1	12.5%	
AZA-2	µg/kg		0.1	12.5%	
AZA-3	µg/kg		0.1	12.5%	
AZA-total	µg AZA-eq./kg		0.1	12.5%	
Free-DTX1	µg/kg		0.1	12.5%	
Free-DTX2	µg/kg		0.1	12.5%	
Free-Okadaic-Acid	µg/kg		0.1	12.5%	
Total-free-OA+DTX-1+DTX-2	µg OA-eq./kg		0.1	12.5%	
Total-DTX1	µg/kg		0.1	12.5%	
Total-DTX2	µg/kg		0.1	12.5%	
Total-Okadaic Acid	µg/kg		0.1	12.5%	
Total-hy-OA+DTX1+DTX2	µg OA-eq./kg		0.1	12.5%	
PTX-1	µg/kg		0.1	12.5%	
PTX-2	µg/kg		0.1	12.5%	
Total-OA-group and PTX-group	µg OA-eq./kg		0.1	12.5%	
YTX	mg/kg		0.1	12.5%	
homo-YTX	mg/kg		0.1	12.5%	
45-OH-YTX	mg/kg		0.1	12.5%	
45-OH-homo-YTX	mg/kg		0.1	12.5%	
Total-YTX-group	mg YTX-eq./kg		0.1	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

Toxicity Equivalency Factors

Total toxicity equivalence for each of the biotoxin groups should be calculated using toxicity equivalency factors (TEFs) as recommended by EFSA, which are contained in the following table:

Toxin Group	Analogue	TEF
OA-group	OA	1
	DTX-1	1
	DTX-2	0.6
AZA-group	AZA-1	1
	AZA-2	1.8

	AZA-3	1.4
PTX-group	PTX-1	1
	PTX-2	1
YTX-group	YTX	1
	homo-YTX	1
	45-OH-YTX	1
	45-OH-homo-YTX	0.5

Analysis

It is advisable to analyse the test materials as soon as possible after receipt. Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery.

One determination of each test material is required, for each determinand. The results of each determinand should be expressed on the test material "as received" i.e. on a wet weight basis. The concentrations should be determined against your own calibration solutions.

If you routinely conduct analyses by more than one technique you may report multiple sets of data.

You should inform the QUASIMEME office staff, who will arrange an additional exercise template file. It is not possible to report two sets of data using different methods on the same exercise template.

The standard solution ampoules contain at least 500 µl, sufficient for 1 injection of a solution into a LC or 1 analyses of the solution by an assay.

Each vial contains sufficient quantity of extract or homogenate for one analysis. The whole transferable contents of each vial should be extracted. To transfer the contents into a preweighed or tared extraction tube, the vial should be fully defrosted, vortex-mixed and the contents poured into the desired container.

Please note all test materials should be stored in the freezer at ca -20°C or less between analyses.

All the results should be reported in µg/kg (YTX-group toxins in mg/kg) on the basis of the wet weight of the test material as provided. The density of the standard solutions are 0.7918 g/ml (MeOH) and the density of the **lipophilic and DSP/AZP extracts** is 0.834 g/ml. For the tissues, the weight of material should be determined prior to analysis.

Please note that if your laboratory does not report on a given analogue, e.g. DTX-1, then your laboratory should not report the sum of toxins, since this will give 2 z-scores out of line and will possibly make data-analysis more difficult for the remaining laboratories which did determine this analyte.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

ROUND	2017 - 1
BT-12	PSP Shellfish Toxins
Test materials	QST230BT, QST231BT, QST232BT and QST233BT

Objective

This study covers the determination of paralytic shellfish toxins in shellfish tissue.

Test Materials and storage

The test materials were supplied by the Marine Institute, Galway, Ireland.

Shellfish tissue test materials are supplied in a plastic 5ml vial, each vial contains sufficient material for one-shot analysis of the paralytic shellfish toxins.

Each batch of material was prepared in bulk. The level of within and between sample homogeneity and stability was determined. All materials have been shown to be homogeneous at or below the intake mass normally used, and stable for the purposes of the test.

Begin the analysis as soon as possible, preferably within 7 days of receipt.

The shellfish tissue test materials should be stored at -20°C, or a lower temperature, immediately upon receipt, until analysis. The test materials have been shipped on cool packs. The cool packs will thaw slowly during transit, but this does not significantly affect the stability of the material for the purpose of this study, providing the test materials themselves are stored as directed, immediately on receipt.

Sample Code	Description
QST230BT	Oyster tissue homogenate
QST231BT	Mussel tissue homogenate
QST232BT	Mussel tissue homogenate
QST233BT	Mussel tissue homogenate

Precaution

Some of the substances may present a health hazard and can be biologically active. Please ensure that your analytical and handling procedures for this material have been fully assessed for safety and that all specified precautions are taken.

Determinands and concentration ranges

The final (total toxicity) result for each test material should be reported as µg STX dihydrochloride equivalents/kg (such that HPLC, MBA and ELISA results are comparable).

Participants using HPLC methods should also report each PSP analogue identified and give individual analogue concentrations in µmol/kg sample.

Participants using HPLC methods should use the specific toxicities as they appear in the [EFSA](#) Scientific Opinion of the Panel on Contaminants in the Food Chain for Marine Biotoxins in shellfish: STX group (see below).

Determinand	Unit	Concentration range	Error		AA-EQS
			Const	Prop	
11-OH-STX	µmol/kg		0.1	12.5%	
C1	µmol/kg		0.1	12.5%	
C1,2	µmol/kg		0.1	12.5%	
C2	µmol/kg		0.1	12.5%	
C3	µmol/kg		0.1	12.5%	
C3,4	µmol/kg		0.1	12.5%	
C4	µmol/kg		0.1	12.5%	
dc-GTX1	µmol/kg		0.1	12.5%	
dc-GTX1,4	µmol/kg		0.1	12.5%	
dc-GTX2	µmol/kg		0.1	12.5%	
dc-GTX2,3	µmol/kg		0.1	12.5%	
dc-GTX3	µmol/kg		0.1	12.5%	
dc-GTX4	µmol/kg		0.1	12.5%	
dc-NEO	µmol/kg		0.1	12.5%	
dc-STX	µmol/kg		0.1	12.5%	
GTX-1	µmol/kg		0.1	12.5%	
GTX-2	µmol/kg		0.1	12.5%	
GTX-3	µmol/kg		0.1	12.5%	
GTX-4	µmol/kg		0.1	12.5%	
GTX-5	µmol/kg		0.1	12.5%	
GTX-6	µmol/kg		0.1	12.5%	
NEO	µmol/kg		0.1	12.5%	
STX	µmol/kg		0.1	12.5%	
Total toxicity	µgSTXdiHCl-eq/kg		2	12.5%	
GTX-2,3	µmol/kg		0.1	12.5%	
GTX-1,4	µmol/kg		0.1	12.5%	

Results should be reported for as many of these determinands as possible. Take this opportunity either to develop your methodology or check your performance on the less common determinands.

TEFs recommended by the EFSA

Determinand	TEF
STX	1
NeoSTX	1
GTX-1	1
GTX-2	0.4
GTX-3	0.6
GTX-4	0.7
GTX-5	0.1
GTX-6	0.1
C2	0.1
C4	0.1
dc-STX	1
dc-NeoSTX	0.4
dc-GTX-2	0.2
dc-GTX-3	0.4

Analysis

It is advisable to analyse the test materials as soon as possible after receipt. Use your normal validated methods and procedures to analyse the test materials. This may include correcting for blanks and for recovery. One result per test material is required, for the STX analogues individually and as total STX-equivalents. The results of each determinand should be expressed on the test material "as received" i.e. on a wet weight basis. The concentrations should be determined against your own calibration solutions.

Each vial contains sufficient quantity of homogenate for one analysis. The whole transferable contents of each vial should be extracted, and one result reported for the sum of the two isomers (if participants are using the AOAC 2005.06 method then those toxins that co-elute (eg GTX1 and GTX-4, GTX-2 and GTX-3, dcGTX-2 and dcGTX-3 or C-1 and C-2) must be reported using the higher toxicity factor of the two isomers. For example if participants find the presence of GTX-1,4 (co-eluting) in the sample then they should report the sum of the two isomers in the GTX-1,4 column. To transfer the contents into a preweighed or tared extraction tube, the vial should be fully defrosted, vortex-mixed and the contents poured into the desired container.

All results should be reported in TEQ values on the basis of wet weight of the test material as provided. The weight of the shellfish tissue test material should be determined prior to analysis.

Reporting

One result for each determinand in each test material should be reported using the Data Submission Forms which are placed on the Participant Site

It is not possible to report two sets of data using different methods in the same Data Submission Forms.

If you routinely conduct analyses by more than one technique you may report multiple sets of data. You should inform the QUASIMEME office staff, who will arrange an additional Data Submission Form.

Please report your Method information together with the results.

Reporting of Results and Analytical Methods

Units

The units of measurement are given in the Data Submission Forms. Ensure that the concentration of each determinand is reported in the units given. This may differ from your normal units for reporting; it is essential that all data reported are comparable. It is not possible for you to alter the units for reporting in the Data Submission Forms.

The precision of the reported results should reflect the level of uncertainty of the measurement in your laboratory

Reporting Left Censored Values

If the concentration of a determinand is below the detection limit of your method, you may wish to report the value as less than the detection limit. To do this, you should report your detection limit, either as a negative number or preceded by the "less-than" symbol, <. I.e. to report a value less than a detection limit of 10, report either "-10" or "<10". The system will identify either of these formats as left censored ("less-than") values. Left censored values are included in the statistical evaluation of the data, and in the reports.

Method Codes

You are kindly asked to report your methods used, by the Method codes given in the Data Submission Forms. When the method used by your laboratory can not be chosen by one of the MIC (Method Information Code) options given in the Data Submission Form, please select others (option Z) and provide us with the details of the method used by your lab.

Return of Data

Upload all analytical data to the QUASIMEME site only with the Data Submission Forms on the Participant Site. This allows a rapid and accurate transfer of your data and an early report to you. Additional information and comments may be provided as attached files.

Only data submitted using the Data Submission Forms can be included in the assessment. Return the results to the WEPAL-QUASIMEME Project Office in Wageningen no later than **1 July 2017**. Data arriving after this deadline may not be entered into the database or appear in the report.

If you have further information on additional methods used or specific ways in which we can improve the data transfer, please inform the QUASIMEME Project Office (Quasimeme@wur.nl). Your co-operation is appreciated and will help the assessors in the data analysis and in providing appropriate advice in case of any analytical difficulties.

Please observe the following guidelines, to reduce the need for additional checks, replies and enquires:

Data should only be submitted to the WEPAL-QUASIMEME Project Office when all quality checks have been made. If data are submitted beyond the deadline, they might not be included in the report. Data submitted after the issue of the report will not be included in the report, and these data will also not be included as part of the consensus value. Any certificate prepared with data submitted late will include the statement "Data submitted after report issued". No data will be re-entered into the database after the report is issued. No data will be changed in the database UNLESS there is evidence that QUASIMEME or data transfer has caused an error. In such cases QUASIMEME will undertake a quality query to investigate the problem and inform the participant of the outcome of the Query.

The assigned values will be calculated based on the assessment of all data returned, using the Cofino model. The report for each study, including each laboratory's individual assessment and z-scores, will be distributed to participants no later than 1 August 2017. Background information on the data assessment will be provided with the reports.

Collusion and Falsification of Results

QUASIMEME accepts that most participants operate with professional integrity and that data returned as part of the LP studies are correct and are submitted without interference or collusion. However, in some

circumstances, data or information may be influenced by, for example, (i) repeated analyses and submitting mean data, or (ii) collaboration with colleagues undertaking the same study.

QUASIMEME checks for evidence of collusion and confirm to all participants that such activity is contrary to professional scientific conduct and will only nullify the benefits of the LP studies to accreditation bodies and analysts alike.

QUASIMEME reserves the right to withdraw participation of any institute who, in the opinion of the Scientific Assessment Group, has submitted data following collusion or falsification. This statement is made as a formal requirement for accreditation for Laboratory Performance Studies under ISO17043.

ANNEX 1 Notification of damaged test materials.

You do not need to notify QUASIMEME if the test materials arrived in good condition

Client number :

Damaged container number :

Loss of weight container number :

I request a new test material for :..... due to :

Date :

Signature :.....

Name of participant :.....

Name and address of institute :

.....

.....

Telephone number :.....

Fax number :.....

Return this form to :

WEPAL-QUASIMEME Project Office

P.O. Box 8005

6700 EC Wageningen

The Netherlands

Fax No : +31(0)317 486 546

E-mail : QUASIMEME@wur.nl

ANNEX 2 Instructions for login into Participant Site

Login to <https://www.participants.wepal.nl> or www.quasimeme.org

Type in your Username and password into the box

Ask the WEPAL-QUASIMEME Project Office when the login information is unknown

Click login

Select the correct program

Enter your results

Lower than results will be automatically transferred into – values.

Click on the save button to store your data into the database

ANNEX 3 Total Lipid Extraction According to Smedes

This method is based on research carried out by Foppe Smedes.

See : Determination of total lipid using non-chlorinated solvents

Smedes, F., Analyst 124 (1999) : 1711-1718.

Instruments and Chemicals

- Balance with a precision of 0.1 mg
- Ultra Turrax
- Centrifuge capable of holding 100 ml tubes or glass jar at a speed of 2000 rpm¹
- Heated waterbath with condensers.
- Evaporation flasks in suitable shape and size
- Pipettes
- Deionised water
- Isopropanol
- Cyclohexane
- Solution of 13 % (w/w) isopropanol in cyclohexane.

Procedure

- Carry out a dry-weight determination on a representative portion of the test material to be analyses.
- Take a portion of wet test material, which does not contain more than 1g lipid or 8g of water.
- Weigh the test material with known moisture content in a 100ml centrifuge tube or appropriate glass jar.
- Add 18ml isopropanol and 20ml cyclohexane.
- Mix with Ultra Turrax for two minutes.
- Add W ml of water. W is calculated by :

$$W = 22 - \frac{\text{Sampleintake (g)} * \text{moisturecontent (\%)}}{100}$$

- Mix with Ultra Turrax for another minute.
- Separate the phases by centrifugation².
- Transfer as much as possible of the organic phase to an evaporation flask (by small pipette). Filtration is optional but makes the method more robust³.

¹ When a centrifuge is not used, the phases may take time to separate and the interface is less sharp, which can result in a low recovery of the organic phase. A check should be made to determine whether > 80% of the organic phase has been recovered (18 ml). A third extraction is recommended in the case of a lower recovery.

² Some tissues, like liver extracts, form an emulsion which can be prevented by replacing the water by 1 M HClO₄ to denature the proteins. The addition of NaCl may also help.

³ In some cases the organic phase may contain some tissue particles when using the B & D Method. This also depends on the mixing method used (e.g. ultra sonic). When this occurs the extract should be filtered by passing the extract through a glass column plugged with ca 2cm of cottonwool which has previously been extracted with solvent.

- Add 20 ml cyclohexane containing 13%(w/w) isopropanol and mix for one minute by Ultra Turrax.
- Centrifuge.
- Transfer the upper phase to the flask containing the first extract and evaporate the solvent.
- Quantitatively transfer the residue to a weighed wide-mouth cup by using a few ml of the cyclohexane/isopropanol mixture or diethylether.
- Evaporate in a moderately warm place to dryness (do not boil). The temperature used should be 5- 10 °C below the boiling point of the washing solvent. Evaporation may be assisted by a stream of nitrogen.
- Further dry the residue for one hour at 105 °C
- Weigh the residue and calculate the lipid content from the intake.