



**WAGENINGEN EVALUATING PROGRAMS
FOR ANALYTICAL LABORATORIES**

Certificate of Analysis

BIMEP

International Biomass Exchange Program

REFERENCE MATERIAL

BIMEP sample 412

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model mean and standard deviation are calculated using all reported data when at least 8 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into three sections: Consensus Values, Indicative Values and Values for Information. The division is made on the reliability of the data. Consensus Values are based on at least 16 results while the coefficient of variation is smaller than 25 %. Indicative Values are based on at least 8 and less than 16 results or a coefficient of variation between 25 % and 50 %. Other values, based on more than 2 and less than 8 results or a coefficient of variation higher than 50 %, are given for information only.

In the sections with Consensus Values and Indicative Values the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median and MAD (Median of Absolute Deviation) and the uncertainty in the consensus values. The confidence limits (at 95 % probability) are calculated for these determinands.

In the section with Information Values the following parameters are given: median, MAD and number of results. For determinands which have at least 5 results reported as smaller than (<) the median of these 'smaller than results' is calculated. In some cases this median of '<' values is much smaller than median and mean of the indicative values. This may be caused by a too optimistic (too low) value for the detection limit reported by a (small) majority of participating laboratories who report '<' -values.

All values, expressed on a weight basis (kg or %), are reported in oven dry (105 °C) material. Moisture is reported in the material as received.

Sample information

WEPAL reference materials are from natural sources only. There is no spiking, mixing or other alterations of the samples. For sample preparation the BIMEP samples are dried at 40 °C and milled to pass a 0.5 mm sieve.

This BIMEP sample 412 of Reed / Phragmites communis from Netherlands is prepared for the WEPAL proficiency programs. The sample is used in 4 periods (or rounds). The results on which the values in this report are based were taken from the periods given in the following table.

Year	Round	Number
2018	3	4
2016	2	2
2011	1	1
2008	3	2

Consensus Values BIMEP 412

Method: General Analysis

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
Ash	% (m/m)	30.4	1.18	3.9	39	30.2	0.79	0.24	29.98	-	30.75
Moisture	% (m/m)	5.39	0.870	16.1	40	5.29	0.595	0.172	5.11	-	5.67
Calorific Value (gross)	MJ/kg	13.0	0.44	3.4	37	13.0	0.30	0.09	12.87	-	13.16
Volatile Matter	% (m/m)	56.5	1.80	3.2	28	57.0	1.26	0.42	55.8	-	57.2

Method: Elementary Analysis

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
Carbon (C)	% (m/m)	33.6	1.20	3.6	37	33.7	0.82	0.25	33.16	-	33.96
Hydrogen (H)	% (m/m)	4.37	0.334	7.6	33	4.39	0.220	0.073	4.25	-	4.49
Nitrogen (N)	% (m/m)	0.840	0.0546	6.5	35	0.825	0.0360	0.0115	0.821	-	0.859
Cl	g/kg	7.31	1.034	14.1	31	7.18	0.743	0.232	6.93	-	7.69
S	g/kg	2.24	0.482	21.5	34	2.22	0.345	0.103	2.07	-	2.40

Method: Minor Elements

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
Mn	mg/kg	123	14.2	11.6	18	126	9.6	4.2	116	-	130
Zn	mg/kg	44.2	3.06	6.9	16	44.0	2.10	0.96	42.6	-	45.8

Indicative Values BIMEP 412

Method: Major Elements

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
Al	g/kg	13.2	2.54	19.3	12	13.5	1.75	0.92	11.6	-	14.8
Ca	g/kg	4.74	1.412	29.8	13	5.10	1.040	0.490	3.89	-	5.58
Fe	g/kg	6.77	0.723	10.7	13	7.00	0.500	0.251	6.34	-	7.20
K	g/kg	15.9	2.42	15.3	13	16.5	1.72	0.84	14.4	-	17.3
Mg	g/kg	1.97	0.343	17.4	13	1.96	0.240	0.119	1.76	-	2.18
Na	g/kg	3.33	0.546	16.4	12	3.40	0.373	0.197	2.99	-	3.67
P	g/kg	1.34	0.416	31.1	12	1.38	0.313	0.150	1.08	-	1.60
Si	g/kg	79.5	26.75	33.6	13	84.1	16.53	9.27	63.5	-	95.6

Method: Minor Elements

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
As	mg/kg	3.03	0.587	19.4	13	3.17	0.430	0.203	2.67	-	3.38
Cd	mg/kg	0.168	0.0750	44.7	8	0.200	0.0620	0.0331	0.107	-	0.229
Co	mg/kg	2.60	0.360	13.8	12	2.55	0.250	0.130	2.38	-	2.83
Cr	mg/kg	16.0	3.74	23.4	14	17.0	2.54	1.25	13.9	-	18.2
Cu	mg/kg	10.2	2.23	21.9	14	10.9	1.60	0.75	8.92	-	11.5
Hg	mg/kg	0.0379	0.0061	16.2	10	0.0375	0.0041	0.0024	0.0336	-	0.0422
Mo	mg/kg	1.30	0.298	22.9	10	1.35	0.200	0.118	1.09	-	1.51
Ni	mg/kg	8.72	1.000	11.5	14	8.80	0.710	0.334	8.15	-	9.29
Pb	mg/kg	9.45	1.614	17.1	12	9.81	1.125	0.582	8.44	-	10.5
Ti	mg/kg	653	239.0	36.6	9	623	163.0	99.6	472	-	833
V	mg/kg	13.8	2.92	21.2	13	14.3	1.90	1.01	12.1	-	15.6



BIMEP

Informative Values BIMEP 412

Method: Elementary Analysis

Element	Unit	Median	MAD	N
Oxygen (O)	% (m/m)	32.7	2.49	3

Method: Water Soluble Elements

Element	Unit	Median	MAD	N
Cl	g/kg	7.13	0.019	3
K	g/kg	11.0	0.16	4
Na	g/kg	1.65	0.105	4

Method: Minor Elements

Element	Unit	Median	MAD	N
Ba	mg/kg	78.3	29.60	12
Be	mg/kg	0.330	0.0750	6
F	mg/kg	66.1	25.50	9
Sb	mg/kg	1.60	1.300	7
Se	mg/kg	1.15	0.090	5
Sn	mg/kg	1.10	0.550	8
Te	mg/kg	-	-	0
Tl	mg/kg	-	-	0

Results smaller than (<)

Median of <	N
0.200	6
0.750	8
0.650	6