



**WAGENINGEN EVALUATING PROGRAMS  
FOR ANALYTICAL LABORATORIES**

---

**Certificate of Analysis**

**BIMEP**

**International Biomass Exchange Program**

**REFERENCE MATERIAL**

**BIMEP sample 417**

---

## General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model mean and standard deviation are calculated using all reported data when at least 8 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into three sections: Consensus Values, Indicative Values and Values for Information. The division is made on the reliability of the data. Consensus Values are based on at least 16 results while the coefficient of variation is smaller than 25 %. Indicative Values are based on at least 8 and less than 16 results or a coefficient of variation between 25 % and 50 %. Other values, based on more than 2 and less than 8 results or a coefficient of variation higher than 50 %, are given for information only.

In the sections with Consensus Values and Indicative Values the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median and MAD (Median of Absolute Deviation) and the uncertainty in the consensus values. The confidence limits (at 95 % probability) are calculated for these determinands.

In the section with Information Values the following parameters are given: median, MAD and number of results. For determinands which have at least 5 results reported as smaller than (<) the median of these 'smaller than results' is calculated. In some cases this median of '<' values is much smaller than median and mean of the indicative values. This may be caused by a too optimistic (too low) value for the detection limit reported by a (small) majority of participating laboratories who report '<' -values.

All values, expressed on a weight basis (kg or %), are reported in oven dry (105 °C) material. Moisture is reported in the material as received.

## Sample information

WEPAL reference materials are from natural sources only. There is no spiking, mixing or other alterations of the samples. For sample preparation the BIMEP samples are dried at 40 °C and milled to pass a 0.5 mm sieve.

This BIMEP sample 417 of Compost from Switzerland is prepared for the WEPAL proficiency programs. The sample is used in 3 periods (or rounds). The results on which the values in this report are based were taken from the periods given in the following table.

Year	Round	Number
2019	1	3
2013	2	2
2009	1	1

**Consensus Values BIMEP 417**

**Method: General Analysis**

<b>Element</b>	<b>Unit</b>	<b>Mean</b>	<b>Std.Dev.</b>	<b>CV %</b>	<b>N</b>	<b>Median</b>	<b>MAD</b>	<b>Uncertainty</b>	<b>95 % confidence limits</b>		
Ash	% (m/m)	71.2	0.89	1.2	29	71.0	0.61	0.21	70.85	-	71.53
Moisture	% (m/m)	3.79	0.258	6.8	29	3.76	0.171	0.060	3.69	-	3.88
Calorific Value (gross)	MJ/kg	5.68	0.338	5.9	22	5.66	0.240	0.090	5.53	-	5.83
Volatile Matter	% (m/m)	26.9	1.25	4.6	23	26.9	0.90	0.33	26.3	-	27.4

**Method: Elementary Analysis**

<b>Element</b>	<b>Unit</b>	<b>Mean</b>	<b>Std.Dev.</b>	<b>CV %</b>	<b>N</b>	<b>Median</b>	<b>MAD</b>	<b>Uncertainty</b>	<b>95 % confidence limits</b>		
Carbon (C)	% (m/m)	16.8	1.01	6.0	25	16.7	0.73	0.25	16.4	-	17.2
Hydrogen (H)	% (m/m)	1.70	0.148	8.7	23	1.71	0.100	0.039	1.64	-	1.77
Nitrogen (N)	% (m/m)	1.42	0.122	8.6	25	1.38	0.090	0.031	1.37	-	1.47

**Method: Minor Elements**

<b>Element</b>	<b>Unit</b>	<b>Mean</b>	<b>Std.Dev.</b>	<b>CV %</b>	<b>N</b>	<b>Median</b>	<b>MAD</b>	<b>Uncertainty</b>	<b>95 % confidence limits</b>		
Mn	mg/kg	634	103.6	16.3	16	631	72.2	32.4	579	-	689

## Indicative Values BIMEP 417

## Method: Elementary Analysis

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
S	g/kg	1.46	0.714	48.8	24	1.51	0.500	0.182	1.16	-	1.76

## Method: Major Elements

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
Al	g/kg	25.4	4.24	16.7	13	26.1	2.83	1.47	22.9	-	28.0
Ca	g/kg	47.5	5.07	10.7	13	48.1	3.46	1.76	44.4	-	50.5
Fe	g/kg	17.3	6.22	36.0	13	18.0	4.60	2.16	13.6	-	21.0
K	g/kg	15.6	1.84	11.8	13	16.5	1.43	0.64	14.5	-	16.7
Mg	g/kg	8.71	0.501	5.8	13	8.72	0.350	0.174	8.41	-	9.01
Na	g/kg	4.23	0.770	18.2	13	4.38	0.510	0.267	3.77	-	4.69
P	g/kg	3.54	0.776	21.9	13	3.74	0.530	0.269	3.08	-	4.01
Si	g/kg	182	71.1	39.1	13	168	49.2	24.6	139	-	224

## Method: Minor Elements

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
As	mg/kg	7.22	1.453	20.1	12	7.09	1.060	0.524	6.31	-	8.14
Ba	mg/kg	150	46.7	31.1	12	155	30.9	16.9	121	-	180
Cd	mg/kg	0.377	0.1259	33.3	10	0.420	0.1050	0.0498	0.289	-	0.466
Co	mg/kg	6.15	0.614	10.0	11	6.17	0.428	0.231	5.74	-	6.55
Cr	mg/kg	109	20.1	18.5	14	108	13.8	6.7	97.0	-	120
Cu	mg/kg	42.6	5.47	12.8	15	43.0	3.61	1.77	39.6	-	45.6
Hg	mg/kg	0.1000	0.0130	13.0	11	0.1012	0.0092	0.0049	0.0913	-	0.109
Mo	mg/kg	3.67	0.623	17.0	9	3.83	0.430	0.260	3.20	-	4.14
Ni	mg/kg	27.0	0.59	2.2	14	27.0	0.45	0.20	26.63	-	27.31
Pb	mg/kg	39.4	6.93	17.6	14	39.2	4.99	2.31	35.5	-	43.4
Sn	mg/kg	3.68	1.391	37.8	10	3.82	0.950	0.550	2.70	-	4.66
V	mg/kg	35.0	7.04	20.1	12	35.5	4.85	2.54	30.6	-	39.5
Zn	mg/kg	135	21.0	15.6	15	129	14.3	6.8	123	-	146



# BIMEP

## Informative Values BIMEP 417

### Method: Elementary Analysis

Element	Unit	Median	MAD	N
Oxygen (O)	% (m/m)	8.50	0.700	5
Cl	g/kg	0.726	0.3631	23

### Method: Water Soluble Elements

Element	Unit	Median	MAD	N
Cl	g/kg	1.58	0.650	3
K	g/kg	3.90	3.093	4
Na	g/kg	0.451	0.2737	4

### Method: Minor Elements

Element	Unit	Median	MAD	N
Be	mg/kg	0.650	0.0550	6
F	mg/kg	64.0	63.99	7
Sb	mg/kg	1.10	0.240	7
Se	mg/kg	0.395	0.1100	4
Te	mg/kg	-	-	0
Ti	mg/kg	1007	654.5	10
Tl	mg/kg	0.250	0.0150	4

### Results smaller than (<)

Median of <	N
1.000	5
0.400	6