



**WAGENINGEN EVALUATING PROGRAMS  
FOR ANALYTICAL LABORATORIES**

---

**Certificate of Analysis**



**International Plant-Analytical Exchange**

**REFERENCE MATERIAL**

**IPE sample 991**

---



## General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model mean and standard deviation are calculated using all reported data when at least 8 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into three sections: Consensus Values, Indicative Values and Values for Information. The division is made on the reliability of the data. Consensus Values are based on at least 16 results while the coefficient of variation is smaller than 25 %. Indicative Values are based on at least 8 and less than 16 results or a coefficient of variation between 25 % and 50 %. Other values, based on more than 2 and less than 8 results or a coefficient of variation higher than 50 %, are given for information only.

In the sections with Consensus Values and Indicative Values the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median and MAD (Median of Absolute Deviation) and the uncertainty in the consensus values. The confidence limits (at 95 % probability) are calculated for these determinands.

In the section with Information Values the following parameters are given: median, MAD and number of results. For determinands which have at least 5 results reported as smaller than (<) the median of these 'smaller than results' is calculated. In some cases this median of '<' values is much smaller than median and mean of the indicative values. This may be caused by a too optimistic (too low) value for the detection limit reported by a (small) majority of participating laboratories who report '<' -values.

All values, expressed on a weight basis (kg or %), are reported in oven dry (105 °C) material. Moisture is reported in the material as received.

## Sample information

WEPAL reference materials are from natural sources only. There is no spiking, mixing or other alterations of the samples. For sample preparation the IPE samples are dried at 70 °C and milled to pass a 0.5 mm sieve.

This IPE sample 991 of Juniper-shoot / Juniperus l. from Netherlands is prepared for the WEPAL proficiency programs. The sample is used in 1 period (or round). The results on which the values in this report are based were taken from the period given in the following table.

Year	Round	Number
1994	3	3



## Consensus Values IPE 991

### Method: Inorganic Chemical Composition

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
B	mg/kg	33.7	3.81	11.3	75	33.9	2.70	0.55	32.9	-	34.6
Ca	g/kg	15.7	1.12	7.1	140	15.7	0.78	0.12	15.56	-	15.94
Cl (as Cl)	g/kg	2.50	0.213	8.5	23	2.52	0.138	0.056	2.41	-	2.59
Cu	mg/kg	4.83	0.770	15.9	122	4.96	0.520	0.087	4.70	-	4.97
Fe	mg/kg	303	49.1	16.2	123	300	34.0	5.5	294	-	311
K	g/kg	11.2	0.73	6.5	147	11.3	0.51	0.08	11.11	-	11.34
Mg	g/kg	1.36	0.086	6.3	140	1.36	0.058	0.009	1.34	-	1.37
Mn	mg/kg	31.6	2.82	8.9	132	31.5	1.95	0.31	31.1	-	32.1
N - Kjeldahl (as N)	g/kg	17.4	0.71	4.1	107	17.5	0.49	0.09	17.26	-	17.53
Na	mg/kg	251	57.4	22.9	92	260	39.3	7.5	239	-	263
Ni	µg/kg	1370	339	24.7	21	1300	210	93	1221	-	1529
P (as P)	g/kg	1.82	0.117	6.4	142	1.83	0.082	0.012	1.81	-	1.84
Pb	µg/kg	6090	1427	23.4	30	6250	1000	326	5550	-	6620
S (as S)	g/kg	1.82	0.179	9.8	49	1.83	0.128	0.032	1.77	-	1.87
Zn	mg/kg	61.6	4.38	7.1	131	61.5	3.04	0.48	60.8	-	62.4

### Method: Real totals

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
N - elementary	g/kg	17.9	1.18	6.6	33	17.9	0.80	0.26	17.5	-	18.4



## Indicative Values IPE 991



### Method: Inorganic Chemical Composition

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
As	µg/kg	192	16.4	8.5	10	192	12.0	6.5	180	-	203
Cd	µg/kg	125	34.9	27.8	33	123	23.0	7.6	113	-	138
Co	µg/kg	185	28.0	15.1	8	185	20.5	12.4	162	-	208
Cr	µg/kg	1010	277	27.5	23	1080	203	72	889	-	1129
Hg	µg/kg	66.9	9.81	14.7	11	70.0	7.00	3.70	60.4	-	73.4
Mo	µg/kg	724	121.8	16.8	9	780	95.0	50.8	632	-	815
N - NO3 (as N)	mg/kg	18.5	4.87	26.3	16	18.2	3.43	1.52	15.9	-	21.1



## Informative Values IPE 991



### Method: Inorganic Chemical Composition

Element	Unit	Median	MAD	N
Ba	mg/kg	15.2	0.90	6
Li	µg/kg	1030	60	4
N - NH4 (as N)	mg/kg	42.5	0.42	3
Sb	µg/kg	266	22.0	5
Se	µg/kg	303	13.0	4
SO4 (as SO4)	g/kg	1.83	0.071	3
Sr	mg/kg	52.0	4.30	7
Ti	mg/kg	10.9	5.16	4
V	µg/kg	1010	110	4