



QUASIMEME

Quality assurance of information
for marine environmental monitoring

Certificate of Analysis



Nutrients in Estuarine and low salinity Seawater

REFERENCE MATERIAL

AQ2 sample 201



Certificate of Analysis AQ2 201

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 10 results while the relative uncertainty is smaller than 6.25%. Indicative Values are based on a relative uncertainty of maximum 35% with at least 4 and less than 10 results or a relative uncertainty higher than 6.25%.

For each determinand the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation) and the uncertainty in the assigned value. The confidence limits (at 95 % probability) are calculated for these determinands.

Sample information

QUASIMEME reference materials cover a range of natural SeaWater species from contaminated waters from the North Sea and/or Mediterranean.

This AQ2 sample 201 of Low salinity seawater unspiked from Baltic Sea is prepared for the QUASIMEME proficiency programs. The results on which the values in this report are based were taken from the periods given in the following table.

Year.Round	Program	Sample Round Id
2022.2	AQ2	QNU388EW



Consensus Values AQ2

Method: Nutrients - AQ2

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits	
Ammonia	µmol/l	1.03	0.247	23.9	30	1.04	0.176	0.056	0.941	- 1.12
Silicate	µmol/l	7.06	0.426	6.0	35	7.11	0.300	0.090	6.92	- 7.21
TOTAL-N	µmol/l	22.5	1.58	7.0	23	22.5	1.09	0.41	21.8	- 23.2



Indicative Values AQ2

Method: Nutrients - AQ2

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits
Nitrite	µmol/l	0.0191	0.0132	68.9	13	0.0270	0.0105	0.0046	0.0112 - 0.0270
Phosphate	µmol/l	0.0334	0.0197	59.1	19	0.0400	0.0150	0.0057	0.0239 - 0.0429
TOTAL-P	µmol/l	0.165	0.0499	30.3	18	0.174	0.0290	0.0147	0.140 - 0.189
TOxN	µmol/l	0.104	0.0682	65.7	15	0.114	0.0460	0.0220	0.0662 - 0.141