



QUASIMEME

Quality assurance of information
for marine environmental monitoring

Certificate of Analysis



Metals in seawater

REFERENCE MATERIAL

AQ3 sample 159



Certificate of Analysis AQ3 159

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 10 results while the relative uncertainty is smaller than 6.25%. Indicative Values are based on a relative uncertainty of maximum 35% with at least 4 and less than 10 results or a relative uncertainty higher than 6.25%.

For each determinand the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation) and the uncertainty in the assigned value. The confidence limits (at 95 % probability) are calculated for these determinands.

Sample information

QUASIMEME reference materials cover a range of natural SeaWater species from contaminated waters from the North Sea and/or Mediterranean.

This AQ3 sample 159 of Seawater (diluted) spiked with metals from North Sea (near Neeltje Jans) is prepared for the QUASIMEME proficiency programs. The results on which the values in this report are based were taken from the periods given in the following table.

Year.Round	Program	Sample Round Id
2020.2	AQ3	QTM305SW



Consensus Values AQ3

Method: Metals - AQ3

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
Arsenic	µg/l	7.35	0.433	5.9	23	7.30	0.300	0.113	7.16	-	7.54
Cadmium	µg/l	0.544	0.0307	5.6	25	0.540	0.0200	0.0077	0.531	-	0.556
Chromium	µg/l	4.51	0.306	6.8	21	4.45	0.203	0.083	4.37	-	4.65
Cobalt	µg/l	1.62	0.206	12.7	17	1.61	0.147	0.063	1.51	-	1.73
Copper	µg/l	4.00	0.475	11.9	22	4.06	0.346	0.127	3.79	-	4.21
Lead	µg/l	0.859	0.0905	10.5	21	0.860	0.0630	0.0247	0.818	-	0.900
Manganese	µg/l	2.23	0.081	3.6	16	2.24	0.057	0.025	2.19	-	2.27
Nickel	µg/l	2.46	0.347	14.1	20	2.49	0.260	0.097	2.30	-	2.62
Vanadium	µg/l	3.29	0.306	9.3	16	3.30	0.211	0.096	3.13	-	3.45
Zinc	µg/l	9.53	0.556	5.8	21	9.49	0.373	0.152	9.27	-	9.78



Indicative Values AQ3

Method: Metals - AQ3

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
Boron	µg/l	1770	275	15.6	9	1850	191	115	1562	-	1977
Iron	µg/l	4.16	0.829	19.9	9	4.15	0.550	0.345	3.54	-	4.79
Silver	µg/l	0.447	0.0647	14.5	9	0.450	0.0440	0.0270	0.398	-	0.496
Tin	µg/l	3.13	0.353	11.3	8	3.05	0.255	0.156	2.84	-	3.41