



QUASIMEME

Quality assurance of information
for marine environmental monitoring

Certificate of Analysis



Sediment

REFERENCE MATERIAL

Sediment sample 66



Certificate of Analysis Sediment 66

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 10 results while the relative uncertainty is smaller than 6.25%. Indicative Values are based on a relative uncertainty of maximum 35% with at least 4 and less than 10 results or a relative uncertainty higher than 6.25%.

For each determinand the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation) and the uncertainty in the assigned value. The confidence limits (at 95 % probability) are calculated for these determinands.

The results of each determinand is expressed on dried sediment.

Sample information

QUASIMEME reference materials cover a range of natural Marine sediment species from contaminated waters from the North Sea and/or Mediterranean. There is no spiking, mixing or other alterations of the samples. For sample preparation the sediment samples are dried at 40 oC and milled to pass a 0.5 mm sieve.

This Sediment sample 66 of mix harbor and open sea sediment from Europoort and Norwegian Trench is prepared for the QUASIMEME proficiency programs. The results on which the values in this report are based were taken from the periods given in the following table.

Year.Round	Program	Sample Round Id
2022.2	MS7	QBC073MS



Indicative Values MS7

Method: Brominated Flame Retardants - MS7

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits
BDE047	µg/kg	0.0541	0.0108	19.9	10	0.0585	0.0073	0.0043	0.0465 - 0.0617
BDE099	µg/kg	0.0479	0.0093	19.4	9	0.0500	0.0065	0.0039	0.0409 - 0.0549
BDE100	µg/kg	0.0135	0.0041	30.1	8	0.0149	0.0033	0.0018	0.0102 - 0.0168
BDE153	µg/kg	0.0175	0.0044	25.1	5	0.0190	0.0028	0.0025	0.0125 - 0.0226
BDE154	µg/kg	0.0132	0.0063	47.6	4	0.0149	0.0042	0.0039	0.0045 - 0.0219
BDE183	µg/kg	0.0196	0.0097	49.3	5	0.0240	0.0074	0.0054	0.0085 - 0.0307
BDE209	µg/kg	0.978	0.2519	25.8	6	1.060	0.1708	0.1285	0.727 - 1.23