



WEPAL-QUASIMEME

Certificate of Analysis



Metals in seawater

REFERENCE MATERIAL

AQ3 sample 202

2026-Aug-11



Certificate of Analysis AQ3 202

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model, the mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 8 results and a maximum relative uncertainty of 6.25%. Indicative Values are based on a maximum relative uncertainty of 35% and a minimum of 4 and maximum of 7 results, or a relative uncertainty greater than 6.25% when there are at least 8 results.

For each measurand, the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation), the uncertainty of the mean (consensus or indicative) value and the relative uncertainty.

Please note: Most WEPAL-QUASIMEME reference materials are found to be stable over the long term (>10 years) for most measurand/matrix combinations. There are a few exceptions known to us as being less stable over the long term. These are organotins in sediment (MS6), ASP in shellfish (BT7), some PAHs and PCBs in sediment (SETOC) and N-NH₄ (as N) & N-NO₃ (as N) in clay soils (ISE).

Sample information

QUASIMEME reference materials cover a range of natural SeaWater species from contaminated waters from the North Sea and/or Mediterranean.

This AQ3 sample 202 of Seawater spiked with metals from North Sea, the Netherlands is prepared for the QUASIMEME proficiency programs. The results on which the values in this report are based were taken from the rounds given in the following table.

Year.Round	Program	Sample Round Id
2026.1	AQ3	QTM392SW



Consensus Values AQ3

Method: Metals - AQ3

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
Copper	µg/l	8.75	1.18	13.5	17	8.90	0.708	0.357	4.08
Cadmium	µg/l	0.401	0.047	11.8	16	0.403	0.021	0.015	3.68
Lead	µg/l	2.07	0.164	7.9	15	2.08	0.088	0.053	2.55
Cobalt	µg/l	1.47	0.139	9.4	11	1.50	0.075	0.052	3.55
Manganese	µg/l	4.76	0.680	14.3	11	4.74	0.320	0.256	5.39
Arsenic	µg/l	3.85	0.399	10.4	13	3.88	0.220	0.138	3.59
Nickel	µg/l	8.08	0.991	12.3	16	8.09	0.658	0.310	3.83
Zinc	µg/l	15.1	1.89	12.5	16	15.0	1.05	0.590	3.92
Boron	µg/l	4356	359	8.2	8	4359	275	159	3.64
Vanadium	µg/l	9.57	1.00	10.5	9	9.63	0.720	0.419	4.37
Tin	µg/l	8.67	0.455	5.3	8	8.63	0.216	0.201	2.32
Uranium	µg/l	3.19	0.090	2.8	8	3.19	0.072	0.040	1.24
Magnesium	mg/l	1193	50.7	4.2	8	1189	29.0	22.4	1.88
Strontium	mg/l	7.07	0.126	1.8	9	7.10	0.070	0.053	0.743



Indicative Values AQ3

Method: Metals - AQ3

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
Iron	µg/l	4.14	1.29	31.1	6	4.14	0.595	0.658	15.9
Chromium	µg/l	2.14	0.440	20.5	12	2.21	0.290	0.159	7.41
Silver	µg/l	0.499	0.036	7.2	5	0.501	0.015	0.020	4.01
Thallium	µg/l	0.821	0.107	13.0	4	0.799	0.054	0.067	8.15