



WEPAL-QUASIMEME

Certificate of Analysis



Halogenated Organics in seawater

REFERENCE MATERIAL

AQ5 sample 119

2026-Aug-13



Certificate of Analysis AQ5 119

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model, the mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 8 results and a maximum relative uncertainty of 6.25%. Indicative Values are based on a maximum relative uncertainty of 35% and a minimum of 4 and maximum of 7 results, or a relative uncertainty greater than 6.25% when there are at least 8 results.

For each measurand, the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation), the uncertainty of the mean (consensus or indicative) value and the relative uncertainty.

Please note: Most WEPAL-QUASIMEME reference materials are found to be stable over the long term (>10 years) for most measurand/matrix combinations. There are a few exceptions known to us as being less stable over the long term. These are organotins in sediment (MS6), ASP in shellfish (BT7), some PAHs and PCBs in sediment (SETOC) and N-NH₄ (as N) & N-NO₃ (as N) in clay soils (ISE).

Sample information

QUASIMEME reference materials cover a range of natural SeaWater species from contaminated waters from the North Sea and/or Mediterranean.

This AQ5 sample 119 of Seawater with spike solution from North Sea, the Netherlands is prepared for the QUASIMEME proficiency programs. The results on which the values in this report are based were taken from the rounds given in the following table.

Year.Round	Program	Sample Round Id
2026.1	AQ5	QOC116SW



Consensus Values AQ5

Method: PCBs&OCP - AQ5

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
a-HCH	ng/l	7.11	0.580	8.1	10	7.12	0.360	0.229	3.22
g-HCH	ng/l	5.33	0.668	12.5	9	5.12	0.361	0.278	5.22
Endrin	ng/l	10.9	0.914	8.4	8	11.2	0.620	0.404	3.70
b-HCH	ng/l	5.77	0.451	7.8	10	5.68	0.245	0.178	3.09
d-HCH	ng/l	5.19	0.402	7.8	8	5.20	0.295	0.178	3.43
PCB138	ng/l	9.26	0.170	1.8	9	9.20	0.100	0.071	0.765



Indicative Values AQ5

Method: PCBs&OCP - AQ5

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
Dieldrin	ng/l	10.0	2.26	22.6	10	10.5	1.34	0.895	8.93
pp'-DDE	ng/l	6.11	1.04	17.0	9	6.35	0.548	0.433	7.09
pp'-DDT	ng/l	6.95	1.20	17.2	10	6.65	0.685	0.473	6.81
pp'-DDD	ng/l	9.02	1.68	18.7	10	9.23	0.717	0.665	7.38
HCB	ng/l	7.20	1.95	27.1	10	7.41	1.25	0.770	10.7
Isodrin	ng/l	10.8	3.42	31.6	7	11.0	1.29	1.62	14.9
Aldrin	ng/l	9.33	3.02	32.3	10	9.82	1.68	1.19	12.8
op'-DDT	ng/l	10.1	2.42	23.9	10	9.95	1.53	0.955	9.47
Trifluralin	ng/l	11.6	2.01	17.3	6	12.1	1.43	1.03	8.85
Endosulphan-I	ng/l	4.33	0.717	16.6	6	4.31	0.422	0.366	8.45
Endosulphan-II	ng/l	5.44	0.389	7.2	6	5.46	0.215	0.199	3.65
Pentachlorobenzene	ng/l	15.1	4.30	28.5	8	15.3	2.27	1.90	12.6
PCB28	ng/l	9.41	2.66	28.2	7	9.76	1.26	1.25	13.3
PCB52	ng/l	7.73	1.51	19.6	9	8.00	0.898	0.631	8.16
PCB101	ng/l	6.32	1.72	27.2	8	6.25	1.20	0.760	12.0
PCB105	ng/l	5.59	0.595	10.7	6	5.78	0.250	0.304	5.44
PCB118	ng/l	9.66	2.08	21.5	8	9.38	1.17	0.919	9.52
PCB153	ng/l	9.91	2.57	25.9	9	10.2	1.20	1.07	10.8
PCB156	ng/l	9.03	0.542	6.0	4	9.05	0.250	0.339	3.75
PCB180	ng/l	5.88	1.24	21.1	8	5.75	0.800	0.550	9.35
Heptachlor	ng/l	4.55	1.36	29.9	8	4.51	0.760	0.602	13.2
Heptachlorepoide	ng/l	7.34	2.03	27.7	6	7.25	1.51	1.04	14.1