



QUASIMEME

Quality assurance of information
for marine environmental monitoring

Certificate of Analysis



Pentachlorophenol in seawater

REFERENCE MATERIAL

AQ7 sample 89



Certificate of Analysis AQ7 89

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model, the mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 8 results and a maximum relative uncertainty of 6.25%. Indicative Values are based on a maximum relative uncertainty of 35% and a minimum of 4 and maximum of 7 results, or a relative uncertainty greater than 6.25% when there are at least 8 results.

For each determinand, the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation), the uncertainty of the mean (consensus or indicative) value and the relative uncertainty.

Please note: Most WEPA-QUASIMEME reference materials are found to be stable over the long term (>10 years) for most determinand/matrix combinations. There are a few exceptions known to us as being less stable over the long term. These are organotins in sediment (MS6), ASP in shellfish (BT7), some PAHs and PCBs in sediment (SETOC) and N-NH₄ (as N) in clay soils (ISE).

Sample information

QUASIMEME reference materials cover a range of natural SeaWater species from contaminated waters from the North Sea and/or Mediterranean.

This AQ7 sample 89 of Seawater with spike solution from North Sea, the Netherlands is prepared for the QUASIMEME proficiency programs. The results on which the values in this report are based were taken from the periods given in the following table.

Year.Round	Program	Sample Round Id
2025.1	AQ7	QPP089SW



Indicative Values AQ7

Method: PCBs&OCP - AQ7

Element

Pentachlorophenol

Unit
ng/l

Mean
1384

Std.Dev.
59.7

CV %
4.3

N
5

Median
1360

MAD
40.0

Uncertainty
33.4

Rel.Uncert. %
2.41