



QUASIMEME

Quality assurance of information
for marine environmental monitoring

Certificate of Analysis



ASP shellfish toxins

REFERENCE MATERIAL

BT7 sample 53



Certificate of Analysis BT7 53

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 10 results while the relative uncertainty is smaller than 6.25%. Indicative Values are based on a relative uncertainty of maximum 35% with at least 4 and less than 10 results or a relative uncertainty higher than 6.25%.

For each determinand the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation) and the uncertainty in the assigned value. The confidence limits (at 95 % probability) are calculated for these determinands.

The results of each determinand is expressed on a wet weight basis.

Sample information

QUASIMEME reference materials cover a range of natural Shellfish toxins species from contaminated waters from the North Sea and/or Mediterranean.

This BT7 sample 53 of Scallop tissue (*Pecten Maximus*) from Marine Institute, Galway, Ireland is prepared for the QUASIMEME proficiency programs. The results on which the values in this report are based were taken from the periods given in the following table.

Year.Round	Program	Sample Round Id
2023.1	BT7	QST332BT
2021.2	BT7	QST305BT
2020.1	BT7	QST278BT
2018.1	BT7	QST242BT
2015.1	BT7	QST184BT



Consensus Values BT7

Method: Toxins(SF) - BT7

Element

Domoic+EpiDomoic

Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	95 % confidence limits		
mg/kg	58.7	7.68	13.1	234	57.7	4.72	0.63	57.7	-	59.7