



**WAGENINGEN EVALUATING PROGRAMS
FOR ANALYTICAL LABORATORIES**

Certificate of Analysis

BIMEP

International Biomass Exchange Program

REFERENCE MATERIAL

BIMEP sample 449

General Information

In this report an overview is given of analytical data for this sample collected in our proficiency testing program. The consensus values are calculated using a robust statistical model. With this NDA model, the mean and standard deviation are calculated using all reported data when at least 4 results are left after removal of reported 'lower than' (<) and 0 (= zero) values. No outliers are removed.

This report is divided into two sections: Consensus Values and Indicative Values. The division is made on the reliability of the data. Consensus Values are based on at least 8 results and a maximum relative uncertainty of 6.25%. Indicative Values are based on a maximum relative uncertainty of 35% and a minimum of 4 and maximum of 7 results, or a relative uncertainty greater than 6.25% when there are at least 8 results.

For each determinand the following parameters are given: mean, standard deviation, coefficient of variation, number of results, median, MAD (Median of Absolute Deviation), the uncertainty of the mean (consensus or indicative) value and the relative uncertainty.

All values, expressed on a weight basis (kg or %), are reported as oven-dried (105°C) material. Moisture is reported in the material as received.

Sample information

WEPAL reference materials are from natural sources only. There is no spiking, mixing or other alterations of the samples. For sample preparation, the BIMEP samples are dried at 40°C and milled to pass a 0.5 mm sieve.

This BIMEP sample 449 of Rosa (plant) / Rosa l., from Netherlands, is prepared for the WEPAL proficiency programs. The sample has been used in 1 period (or round). The results on which the values in this report are based were taken from the period given in the following table:

Year	Round	Number
2012	2	4

Consensus Values BIMEP 449

Method: General Analysis

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
Calorific Value (gross)	MJ/kg	19.2	0.150	0.8	10	19.2	0.095	0.059	0.310
Ash	% (m/m)	4.46	0.225	5.0	10	4.46	0.140	0.089	1.99
Moisture	% (m/m)	6.73	0.279	4.1	10	6.72	0.155	0.110	1.64
Volatile Matter	% (m/m)	75.8	1.05	1.4	9	75.7	0.570	0.437	0.576

Method: Elementary Analysis

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
S	g/kg	1.12	0.155	13.8	10	1.13	0.105	0.061	5.45
Carbon (C)	% (m/m)	48.2	0.927	1.9	9	48.0	0.690	0.386	0.802
Hydrogen (H)	% (m/m)	5.95	0.199	3.3	9	5.98	0.120	0.083	1.39
Nitrogen (N)	% (m/m)	1.77	0.073	4.1	9	1.75	0.040	0.030	1.71

Indicative Values BIMEP 449

Method: Elementary Analysis

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
Cl	g/kg	0.512	0.101	19.8	9	0.550	0.068	0.042	8.25

Method: Major Elements

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
Ca	g/kg	6.50	2.14	32.9	5	7.21	1.41	1.19	18.4
K	g/kg	11.8	3.39	28.7	4	12.6	1.81	2.12	17.9
Mg	g/kg	1.82	0.363	20.0	5	1.90	0.230	0.203	11.2
Na	g/kg	0.197	0.034	17.3	5	0.201	0.019	0.019	9.68
P	g/kg	3.23	0.811	25.1	5	3.31	0.510	0.453	14.1
Si	g/kg	1.38	0.738	53.6	5	1.58	0.450	0.412	30.0

Method: Minor Elements

Element	Unit	Mean	Std.Dev.	CV %	N	Median	MAD	Uncertainty	Rel.Uncert. %
Ba	mg/kg	5.26	2.51	47.8	4	5.00	1.25	1.57	29.9
Mn	mg/kg	27.1	3.93	14.5	4	26.9	2.10	2.46	9.07
Zn	mg/kg	59.4	15.0	25.2	4	59.3	8.16	9.37	15.8